

臺北榮民總醫院藥學部

臨床試驗藥局藥品冷藏/冷凍/室溫溫度管制表 106年 4月

	一號冰箱		二號冰箱		三號冰箱		四號冰箱		負20度冰箱		負80度冰箱		室溫	
序號	0520-362		0520-293		1504-0305		1611-0052		1213-0010		1215-0006		1213-0007	
備用	0580-0835		0520-292		1504-0339		1611-0062		1405-0193		1330-0064		1405-0194	
範圍	2~8℃		2~8℃		2~8℃		2~8℃		負15~負25℃		≤負70℃		15~25℃	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2017/4/1	5	5.8	3.6	5.1	5.3	5.6	4.1	4.7	-20.6	-20.2	-81.1	-79.7	19.2	19.8
2017/4/2	5.1	5.7	3.6	5.1	5.3	5.6	4.1	4.7	-20.7	-20.3	-81.2	-79.9	18.8	19.2
2017/4/3	4.9	5.7	3.7	5.2	5.3	5.6	4.1	4.8	-20.7	-20.3	-81.2	-79.9	18.4	18.9
2017/4/4	4.9	5.7	3.7	5.2	5.3	5.6	4.1	4.7	-20.7	-20.3	-81.2	-79.9	18.2	18.9
2017/4/5	4.9	5.6	3.6	5.3	4.3	5.4	4.4	5.4	-20.4	-19.3	-81.2	-79.9	18	19.8
2017/4/6	4.7	5.8	3.9	5.4	4.3	5.5	4.4	5.2	-20.4	-19.9	-81.2	-79.9	18	19.3
2017/4/7	4.9	5.9	3.8	5.3	4.3	5.4	4.6	5.2	-20.4	-19.8	-81.1	-79.8	18.1	20.8
2017/4/8	5.3	5.9	4.1	5.2	4.9	5.3	4.6	5.1	-20.2	-19.9	-81.1	-79.9	18.4	18.9
2017/4/9	5.3	5.9	4.1	5.3	4.9	5.4	4.6	5.1	-20.3	-19.9	-81.2	-79.9	18.3	18.6
2017/4/10	5.3	5.9	3.7	5.3	4.4	5.3	4.6	5.3	-20.4	-19.9	-81.2	-79.9	18.4	20.1
2017/4/11	5.2	5.9	3.6	5.2	4.3	5.6	4.8	5.8	-20.7	-19.9	-81.2	-79.9	18.8	19.9
2017/4/12	5.2	5.9	3.6	5.1	4.3	5.6	4.7	5.3	-20.7	-20.4	-81.1	-79.9	18.7	19.9
2017/4/13	5.3	5.9	3.5	5.1	4.3	5.4	4.7	5.3	-20.8	-20.2	-81.2	-79.9	18.6	20.1
2017/4/14	5.1	6.1	3.5	5.1	4.6	5.6	4.6	5.2	-20.7	-19.7	-81.2	-79.9	18.6	20
2017/4/15	5.1	5.8	3.8	5.1	5.2	5.6	4.6	5.1	-20.4	-19.9	-81.2	-79.9	18.7	18.9
2017/4/16	5.1	5.9	3.8	5.2	5.2	5.5	4.6	5.1	-20.4	-20.1	-81.2	-79.8	18.7	18.8
2017/4/17	5.3	6.1	3.5	5.1	4.3	5.4	4.6	5.1	-20.4	-19.8	-81.2	-79.9	18.6	20
2017/4/18	5.1	5.9	3.5	5.1	4.3	5.3	4.5	5.1	-20.2	-19.8	-81.2	-79.9	18.8	20.3
2017/4/19	5.1	5.8	3.5	5	4.3	5.5	4.6	5.4	-20.3	-19.9	-81.1	-79.9	19	20.1
2017/4/20	4.9	5.9	3.4	5.1	4.4	5.6	4.8	5.6	-20.3	-19.6	-81.2	-79.8	19	20.9
2017/4/21	5.1	5.9	3.6	5.1	4.3	5.6	4.7	5.4	-20.3	-19.5	-81.1	-79.8	18.8	20.6
2017/4/22	5	5.8	3.7	5.1	5.1	5.5	4.7	5.3	-20.3	-19.9	-81.1	-79.9	19.2	19.9
2017/4/23	5.2	5.9	3.7	5.2	5.1	5.4	4.7	5.3	-20.3	-20	-81.2	-79.9	18.8	19.2
2017/4/24	5.1	5.8	3.6	5.1	4.4	5.6	4.4	5.3	-20.4	-18.4	-81.1	-79.9	18.7	20.2
2017/4/25	4.9	5.9	3.4	5.2	4.4	5.7	4.4	5.1	-20.4	-18.3	-81.2	-79.8	19.1	20.7
2017/4/26	5.2	5.9	3.6	5.2	4.4	5.6	4.4	5.3	-20.4	-19.9	-81	-79.8	19.2	20.8
2017/4/27	5.2	5.9	3.5	5.2	4.4	5.6	4.4	5.1	-20.4	-19.9	-81.1	-79.9	19.2	20.8
2017/4/28	5.1	5.9	3.5	5.1	4.4	5.6	4.4	5.1	-20.4	-19.9	-81	-79.9	19	20.2
2017/4/29	5.1	5.9	3.5	5.1	4.5	5.7	4.4	5	-20.4	-20.2	-81.2	-79.9	18.7	19.2
2017/4/30	5.1	5.8	3.5	5.1	5.3	5.6	4.4	5.1	-20.4	-19.9	-81.1	-79.9	18.7	18.8

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師:

藥學部 邱保祥 106
師(三)級藥師 0501

藥學部 溫雅如 106
師(三)級藥師 0501