

臺北榮民總醫院藥學部

臨床試驗藥局藥品冷藏/冷凍/室溫溫度管制表 106年 2月

序號	一號冰箱		二號冰箱		三號冰箱		四號冰箱		負20度冰箱		負80度冰箱		室溫	
	0520-362	0520-293	1504-0305	1611-0052	1213-0010	1215-0006	1213-0007							
備用	0580-0835		0520-292		1504-0339		1611-0062		1405-0193		1330-0064		1405-0194	
範圍	2~8°C		2~8°C		2~8°C		2~8°C		負15~負25°C		≤負70°C		15~25°C	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2017/2/1	5.1	5.9	3.9	5.5	4.7	5.5	3.1	4.4	-21	-18.7	-80.9	-79.8	19.2	20.2
2017/2/2	4.8	5.9	3.8	5.5	4.7	5.5	2.9	4.7	-21.1	-19.9	-80.9	-79.8	19	20.4
2017/2/3	4.7	5.7	3.7	5.4	4.7	5.6	3.1	5.4	-20.9	-20.4	-80.9	-79.8	19.3	20.8
2017/2/4	4.9	5.7	4	5.4	5.1	5.6	4.9	5.6	-20.9	-20.4	-80.9	-79.8	19.5	20
2017/2/5	4.9	5.7	4	5.4	5.1	5.5	4.9	5.6	-20.9	-20.4	-80.9	-79.8	19.4	19.6
2017/2/6	4.9	5.7	3.8	5.4	4.6	5.5	4.9	5.6	-20.9	-20.2	-80.9	-79.9	19.3	20.8
2017/2/7	4.9	5.8	3.9	5.4	4.4	5.3	4.9	5.6	-20.9	-19.9	-80.9	-79.8	19.3	20.9
2017/2/8	5.2	6	3.9	5.4	4.4	5.4	5	5.6	-20.4	-20.2	-80.9	-79.8	19.6	21.1
2017/2/9	5.1	5.9	3.8	5.3	4.4	5.3	5.1	5.6	-20.6	-19.8	-80.9	-79.8	19.4	21.2
2017/2/10	5.1	5.9	3.7	5.6	4.4	5.3	4.9	5.5	-20.8	-20.1	-80.9	-79.8	18.9	20
2017/2/11	5.3	5.9	4	5.3	4.7	5.3	4.9	5.4	-20.9	-20.5	-80.9	-79.8	18.6	19.2
2017/2/12	5.2	5.9	4.1	5.3	4.7	5.3	4.9	5.4	-20.8	-20.4	-80.9	-79.8	18.9	19.1
2017/2/13	5.1	6	3.6	5.3	4.4	5.3	4.9	5.8	-20.9	-20.3	-80.9	-79.9	19.1	20.8
2017/2/14	5.3	6	3.7	5.2	4.6	5.4	5.3	5.9	-20.9	-19.9	-80.9	-79.8	19.6	21.4
2017/2/15	5.3	6	3.9	5.3	4.6	5.4	5.3	5.8	-20.4	-19.7	-80.9	-79.7	20.4	22.2
2017/2/16	5.1	6.1	3.9	5.4	4.6	5.4	5.3	5.8	-20.3	-19.7	-80.8	-79.7	19.5	21.4
2017/2/17	5.1	6	4.1	5.3	4.7	5.4	5.3	5.6	-20.3	-19.9	-80.8	-79.7	19.2	20.3
2017/2/18	5.4	6	3.9	5.2	4.9	5.4	5.3	5.6	-20.3	-19.9	-80.9	-79.7	19	20.2
2017/2/19	5.3	5.9	4	5.2	4.9	5.3	5.2	5.6	-20.4	-19.9	-80.9	-79.8	18.7	19
2017/2/20	5	5.9	3.9	5.2	4.6	5.3	5.2	5.6	-20.3	-19.7	-80.8	-79.6	18.6	19.7
2017/2/21	5.1	5.9	3.8	5.2	4.7	5.4	5.1	5.7	-20.6	-19.8	-80.8	-79.7	18.6	21.1
2017/2/22	4.9	5.9	3.9	5.2	4.7	5.5	5.1	6.1	-20.7	-20	-80.9	-79.7	19	20.3
2017/2/23	5.1	5.9	3.9	5.3	4.7	5.6	5.1	5.8	-20.6	-20.2	-80.9	-79.7	18.8	21.3
2017/2/24	5	5.9	3.9	5.3	4.7	5.5	4.9	5.9	-20.9	-19.9	-80.8	-79.8	18.4	19.6
2017/2/25	5.3	5.9	3.9	5.3	5.2	5.6	5.4	5.9	-20.8	-20.4	-80.9	-79.7	18.3	19.2
2017/2/26	5.2	5.9	3.9	5.2	5.2	5.5	5.4	5.8	-20.8	-20.4	-80.8	-79.8	18.7	19.1
2017/2/27	5.1	5.9	3.9	5.2	5.2	5.6	5.4	5.8	-20.8	-20.4	-80.8	-79.8	18.4	20.3
2017/2/28	4.9	5.9	3.8	5.2	4.7	5.6	5.3	5.8	-20.9	-19.9	-80.8	-79.7	19.1	20.8

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師:

藥學部 邱保祥 106
師(三)級藥師 0301

藥學部 陳盈如 106
師(三)級藥師 0301