

臺北榮民總醫院藥學部

臨床試驗藥局藥品冷藏/冷凍/室溫溫度管制表 106年 1月

| | 一號冰箱 | | 二號冰箱 | | 三號冰箱 | | 四號冰箱 | | 負20度冰箱 | | 負80度冰箱 | | 室溫 | |
|-----------|-----------|-----|----------|-----|-----------|-----|-----------|-----|-----------|-------|-----------|-------|-----------|------|
| 序號 | 0520-362 | | 0520-293 | | 1504-0305 | | 1611-0052 | | 1213-0010 | | 1215-0006 | | 1213-0007 | |
| 備用 | 0580-0835 | | 0520-292 | | 1504-0339 | | 1611-0062 | | 1405-0193 | | 1330-0064 | | 1405-0194 | |
| 範圍 | 2~8°C | | 2~8°C | | 2~8°C | | 2~8°C | | 負15~負25°C | | ≤負70°C | | 15~25°C | |
| 日期 | Min | Max | Min | Max | Min | Max | Min | Max | Min | Max | Min | Max | Min | Max |
| 2017/1/1 | 5.3 | 5.9 | 4.9 | 5.7 | 5.4 | 5.7 | 3.6 | 4.4 | -20.1 | -19.8 | -80.9 | -79.8 | 18.7 | 18.8 |
| 2017/1/2 | 5.3 | 5.9 | 4.9 | 5.7 | 5.4 | 5.8 | 3.6 | 4.4 | -20.1 | -19.8 | -80.9 | -79.9 | 18.7 | 18.8 |
| 2017/1/3 | 5.1 | 6.1 | 4.6 | 5.8 | 4.9 | 5.8 | 3.6 | 5.6 | -20.9 | -19.8 | -80.9 | -79.7 | 18.7 | 19.9 |
| 2017/1/4 | 5.1 | 5.9 | 4.5 | 5.9 | 5.1 | 5.8 | 3.6 | 4.8 | -20.8 | -19.6 | -80.9 | -79.8 | 17.9 | 19.6 |
| 2017/1/5 | 5 | 5.7 | 4.6 | 5.8 | 4.5 | 5.6 | 3.7 | 4.4 | -20.4 | -20 | -80.9 | -79.9 | 17.5 | 18.8 |
| 2017/1/6 | 4.9 | 5.7 | 4.6 | 5.8 | 4.5 | 5.3 | 3.8 | 4.4 | -20.6 | -20 | -80.9 | -79.9 | 17.4 | 18.6 |
| 2017/1/7 | 4.9 | 5.7 | 4.6 | 5.5 | 5.1 | 5.4 | 3.8 | 4.4 | -20.6 | -20.3 | -80.9 | -79.8 | 17.4 | 17.8 |
| 2017/1/8 | 5.2 | 5.8 | 4.6 | 5.4 | 5.2 | 5.6 | 3.7 | 4.4 | -20.4 | -19.9 | -80.9 | -79.8 | 17.9 | 18.8 |
| 2017/1/9 | 5.1 | 5.9 | 4.4 | 5.9 | 4.4 | 5.6 | 3.7 | 4.4 | -20.2 | -19.3 | -80.9 | -79.8 | 18.7 | 21.1 |
| 2017/1/10 | 5.1 | 5.9 | 4.4 | 5.8 | 4.4 | 5.5 | 3.5 | 4.8 | -19.8 | -19.2 | -80.9 | -79.8 | 19.5 | 20.6 |
| 2017/1/11 | 4.9 | 5.8 | 4.4 | 6.2 | 4.5 | 5.3 | 3.6 | 4.4 | -20.2 | -19.4 | -80.9 | -79.8 | 18.3 | 20 |
| 2017/1/12 | 5 | 5.8 | 4.7 | 6.2 | 4.4 | 5.3 | 3.6 | 4.4 | -20.3 | -19.5 | -80.9 | -79.8 | 18.2 | 19.8 |
| 2017/1/13 | 4.8 | 6.1 | 4.9 | 6.2 | 3.9 | 5.2 | 3.5 | 4.4 | -20.4 | -19.6 | -80.9 | -79.8 | 18.6 | 19.7 |
| 2017/1/14 | 5.1 | 5.8 | 4.9 | 5.7 | 4.3 | 4.7 | 3.5 | 4.4 | -20.3 | -19.9 | -80.9 | -79.8 | 18.7 | 19.3 |
| 2017/1/15 | 5.1 | 5.8 | 4.9 | 5.7 | 4.2 | 4.6 | 3.5 | 4.4 | -20.2 | -19.9 | -80.9 | -79.8 | 19.3 | 19.4 |
| 2017/1/16 | 4.8 | 5.9 | 4.7 | 6.5 | 4.1 | 4.7 | 3.3 | 4.4 | -20.2 | -18.3 | -80.9 | -79.8 | 19.3 | 20.8 |
| 2017/1/17 | 5.1 | 5.9 | 4.7 | 6.2 | 3.9 | 4.7 | 3.4 | 4.6 | -20.8 | -19.9 | -80.9 | -79.8 | 19.7 | 21.3 |
| 2017/1/18 | 5.1 | 5.9 | 4.7 | 6.3 | 4 | 4.7 | 3.8 | 4.5 | -20.8 | -19.2 | -80.9 | -79.7 | 20.1 | 21.4 |
| 2017/1/19 | 5 | 5.8 | 4.7 | 6.1 | 4 | 5.1 | 3.1 | 4.5 | -20.2 | -19.2 | -80.9 | -79.8 | 19.3 | 20.8 |
| 2017/1/20 | 5.1 | 5.9 | 4.6 | 6.1 | 4.2 | 5.8 | 3.1 | 4.3 | -20.2 | -19.3 | -80.9 | -79.7 | 19.7 | 21.8 |
| 2017/1/21 | 5.1 | 5.9 | 4.7 | 5.5 | 5.2 | 5.6 | 3.2 | 4.3 | -20.3 | -19.9 | -80.9 | -79.7 | 19.6 | 20.2 |
| 2017/1/22 | 5.2 | 5.8 | 4.7 | 5.5 | 5.2 | 5.5 | 3.2 | 4.3 | -20.7 | -19.9 | -80.9 | -79.8 | 18.2 | 19.6 |
| 2017/1/23 | 4.9 | 6.5 | 4.6 | 6.1 | 4.4 | 5.4 | 3.2 | 4.4 | -20.9 | -19.9 | -80.9 | -79.8 | 17.9 | 19.4 |
| 2017/1/24 | 4.9 | 5.8 | 4.5 | 5.9 | 4.4 | 5.4 | 3.1 | 4.4 | -21.2 | -20.2 | -80.9 | -79.9 | 17.9 | 19.6 |
| 2017/1/25 | 4.8 | 5.8 | 4.5 | 6 | 4.4 | 5.4 | 3.1 | 4.4 | -21.2 | -20.8 | -80.9 | -79.8 | 17.9 | 19.2 |
| 2017/1/26 | 5.1 | 5.8 | 4.4 | 5.9 | 4.6 | 5.4 | 3.2 | 4.5 | -21.4 | -20.9 | -80.9 | -79.9 | 17.7 | 18.9 |
| 2017/1/27 | 5.1 | 5.8 | 4.4 | 5.5 | 5.1 | 5.4 | 3.4 | 4.6 | -21.4 | -20.9 | -80.9 | -79.9 | 17.7 | 17.9 |
| 2017/1/28 | 5.1 | 5.8 | 4.4 | 5.4 | 5.1 | 5.4 | 3.4 | 4.6 | -21.4 | -21 | -80.9 | -79.9 | 17.7 | 18.1 |
| 2017/1/29 | 5.3 | 5.9 | 4.4 | 5.5 | 5.1 | 5.6 | 3.3 | 4.6 | -21.4 | -20.9 | -80.9 | -79.8 | 18.1 | 19.2 |
| 2017/1/30 | 5.2 | 5.9 | 4.4 | 5.5 | 5.2 | 5.5 | 3.4 | 4.5 | -21.4 | -20.8 | -80.9 | -79.8 | 19.2 | 19.3 |
| 2017/1/31 | 5.2 | 5.9 | 4.3 | 5.4 | 5.2 | 5.5 | 3.4 | 4.6 | -21.4 | -20.9 | -80.9 | -79.8 | 19 | 19.2 |

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師: 藥學部溫雅如 106
師(三)級藥師 0201

藥學部陳盈如 0201
師(三)級藥師 0920