

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 112年 06 月

	一號冰箱		二號冰箱		三號冰箱		負20度冰箱		負80度冰箱		室溫1	
序號	0520-362		0520-292		1504-0305		1405-0193		1215-0006		1213-0007	
備用	0850-0835		0520-293		1504-0339		1213-0010		1330-0064		1405-0194	
範圍	2~8℃		2~8℃		2~8℃		負15~負25℃		≤負70℃		15~25℃	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2023/6/1	4.2	5.1	3.2	4.3	3.8	4.4	-20.6	-19.2	-80.3	-78.6	18.8	19.2
2023/6/2	4.3	5.4	3.1	4.3	3.9	4.8	-20.7	-19.9	-80.4	-77.7	18.6	19.8
2023/6/3	4.9	5.4	3.1	4.1	3.8	4.3	-20.6	-20.2	-80.4	-78.5	19.2	19.3
2023/6/4	4.8	5.4	3.1	4.1	3.8	4.3	-20.6	-20.3	-80.4	-78.6	19.2	19.2
2023/6/5	4.7	5.5	3.1	4.2	3.8	4.4	-20.9	-20.2	-80.4	-78.8	19.2	19.7
2023/6/6	4.7	5.4	3.2	4.3	3.8	4.3	-20.9	-19	-80.3	-78.9	19.2	19.8
2023/6/7	4.8	5.6	3.4	4.4	3.8	4.7	-20.1	-18.8	-80.3	-77.6	19.4	19.9
2023/6/8	4.7	5.7	3.8	4.6	3.7	4.8	-20.1	-19.6	-79.7	-78.1	19.1	19.7
2023/6/9	4.7	5.7	4.1	4.7	3.7	5.1	-20.4	-19.6	-79.8	-77.9	18.9	20.2
2023/6/10	4.8	5.5	4.1	4.4	3.9	4.4	-20.4	-20.2	-79.8	-78.8	18.6	18.9
2023/6/11	4.8	5.6	4.1	4.4	3.9	4.4	-20.4	-20.2	-79.8	-78.2	18.6	18.7
2023/6/12	4.7	5.6	3.7	5.1	3.9	4.4	-20.7	-19.9	-79.8	-78.6	18.6	19.2
2023/6/13	4.7	5.8	3.7	4.8	3.9	4.4	-20.8	-20.3	-79.8	-78.6	18.6	19.1
2023/6/14	4.7	5.7	3.7	4.7	3.9	4.5	-20.9	-20.2	-79.8	-77.6	18.6	19.1
2023/6/15	4.7	5.6	3.9	4.6	3.9	4.5	-20.9	-20.2	-79.7	-77.3	18.6	19.2
2023/6/16	4.6	5.4	3.9	4.4	3.9	4.4	-21.2	-19.6	-79.7	-77.4	18.6	19.2
2023/6/17	4.6	5.4	4	4.4	3.9	4.4	-21.2	-19.6	-79.7	-78.2	18.6	19.1
2023/6/18	4.7	5.4	4	4.4	3.9	4.4	-21.1	-20.2	-79.8	-78.8	18.6	18.7
2023/6/19	4.7	5.8	3.8	4.4	3.9	4.6	-21.1	-20.6	-79.8	-78.6	18.4	19
2023/6/20	4.8	5.7	3.8	4.3	3.9	4.7	-21.1	-20.6	-79.7	-78.8	18.6	19.2
2023/6/21	4.7	5.5	3.6	4.4	3.9	4.5	-21.1	-20.5	-79.7	-77.9	18.7	19.3
2023/6/22	4.8	5.4	3.7	4.3	4.1	4.5	-21.1	-20.8	-79.7	-77.9	18.7	18.9
2023/6/23	4.8	5.3	3.7	4.3	4.1	4.5	-21.1	-20.7	-79.6	-77.9	18.6	18.7
2023/6/24	4.8	5.3	3.7	4.3	4.1	4.5	-21.1	-20.8	-79.7	-77.9	18.5	18.6
2023/6/25	4.8	5.4	3.7	4.3	4	4.5	-21.1	-20.8	-79.8	-78.5	18.4	18.5
2023/6/26	4.7	5.3	3.5	4.4	4.1	4.5	-21.1	-20.7	-79.7	-78.6	18.3	18.9
2023/6/27	4.7	5.8	3.5	4.3	4	4.5	-21.1	-19.9	-80.3	-78.1	18.6	19
2023/6/28	4.7	5.4	3.5	4.3	3.9	4.4	-20.9	-20.3	-79.7	-77.8	18.6	19.2
2023/6/29	4.9	5.5	3.8	4.9	4	4.4	-20.9	-20.6	-80.1	-70.4	18.7	19.2
2023/6/30	4.9	5.6	3.6	4.5	4	4.7	-20.9	-19.8	-79.9	-77.9	18.8	19.2

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師: 藥學部 邱保祥
藥學部 邱予芹

總藥師: 藥學部 廖志峰

業管主管: 藥學部主任 周千澄