

臺北榮民總醫院藥學部

臨床試驗藥局細胞毒性藥品冷藏/室溫溫度管制表 112 年 06 月

	六號冰箱		七號冰箱		室溫	
序號	1436-0086		2123-0026		1118-0011	
備用	0436-049		2123-0027		0742-0169	
範圍	2~8℃		2~8℃		15~25℃	
日期	Min	Max	Min	Max	Min	Max
2023/6/1	4.2	5.7	4	4.6	20.2	20.6
2023/6/2	4.2	4.9	4.1	4.7	20.1	20.6
2023/6/3	4.2	4.9	4.1	4.7	20.4	21.3
2023/6/4	4.2	4.9	4.1	4.6	20.3	20.6
2023/6/5	4.2	4.9	4.1	4.7	20.3	20.9
2023/6/6	4.2	4.9	4	4.7	20.4	21.4
2023/6/7	4.2	4.9	4.1	4.7	20.7	21.3
2023/6/8	4.2	4.9	3.9	4.7	20.6	21.2
2023/6/9	4.2	4.9	3.7	4.6	20.6	21.7
2023/6/10	4.2	4.9	3.7	4.4	20.5	21.3
2023/6/11	4.2	4.9	3.7	4.4	20.3	20.8
2023/6/12	4.1	4.9	3.7	4.5	20.3	20.8
2023/6/13	4.1	4.9	3.8	4.5	20.3	20.4
2023/6/14	4	4.8	3.7	4.7	20.3	20.8
2023/6/15	4.1	4.9	3.8	4.6	20.1	20.6
2023/6/16	4.3	5.2	3.7	4.6	19.9	20.3
2023/6/17	4.5	5.2	3.7	4.5	19.7	20.1
2023/6/18	4.5	5.3	3.7	4.5	19.6	19.9
2023/6/19	4.3	5.2	3.7	4.5	19.5	19.9
2023/6/20	4.3	5.1	3.7	4.5	19.6	20.2
2023/6/21	4.3	5.1	3.8	4.6	19.7	20.5
2023/6/22	4.3	5.1	3.8	4.6	19.9	20.3
2023/6/23	4.3	5.1	3.8	4.6	19.8	20.1
2023/6/24	4.3	5.1	3.8	4.6	19.7	19.9
2023/6/25	4.3	5.1	3.8	4.6	19.5	19.7
2023/6/26	4.2	5.1	3.8	4.6	19.5	20.1
2023/6/27	4.1	4.9	3.7	4.6	19.7	20.1
2023/6/28	4.1	4.9	3.7	4.5	19.8	20.1
2023/6/29	4.2	4.9	3.7	4.5	19.9	20.4
2023/6/30	4.1	4.9	3.7	4.5	20.1	20.7

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師: 藥學部 邱保祥 112
藥學部 邱予芹 112

總藥師: 藥學部 廖志峰 112
藥學部 廖志峰 112

業管主管: 藥學部 周千澄 112
科主任 周千澄 112