

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 112年 07 月

	一號冰箱		二號冰箱		三號冰箱		負20度冰箱		負80度冰箱		室溫1	
序號	0520-362		0520-292		1504-0305		1405-0193		1215-0006		1213-0007	
備用	0850-0835		0520-293		1504-0339		1213-0010		1330-0064		1405-0194	
範圍	2~8℃		2~8℃		2~8℃		負15~負25℃		≤負70℃		15~25℃	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2023/7/1	4.9	5.5	3.7	4.2	4.3	4.7	-22.5	-22.2	-79.9	-77.8	18.8	19
2023/7/2	4.9	5.5	3.7	4.2	4.3	4.7	-22.6	-22.2	-79.9	-77.9	18.7	18.8
2023/7/3	5.1	5.6	3.7	4.2	3.9	4.6	-22.5	-21.2	-79.8	-78.9	18.5	19.1
2023/7/4	4.8	5.7	3.4	4.2	3.6	4.4	-22.4	-21.2	-79.9	-78.9	18.7	19.2
2023/7/5	4.8	5.9	3.4	4.3	3.4	4.1	-22.5	-22.1	-81.3	-77.4	18.8	19.2
2023/7/6	4.4	5.3	3.3	4.2	3.5	4.2	-22.5	-21.3	-81.3	-80.9	18.8	19.5
2023/7/7	4.3	5.1	3.3	4.2	3.6	4.2	-21.6	-20.9	-81.9	-74.1	18.8	19.2
2023/7/8	4.4	5.1	3.3	3.9	3.6	4.1	-21.6	-21.4	-82	-81	18.8	18.9
2023/7/9	4.5	5	3.3	3.9	3.6	4.1	-21.6	-21.4	-82	-81.1	18.7	18.8
2023/7/10	4.4	5.2	3.3	4.9	3.6	4.3	-21.7	-21.2	-82	-81.1	18.6	19.3
2023/7/11	3.1	5.9	3.4	4.5	3.7	4.3	-21.9	-21.5	-82	-80.7	18.8	19.4
2023/7/12	3	5.3	3.2	4.1	3.8	4.4	-21.9	-21.2	-82	-81.2	18.9	19.6
2023/7/13	3.8	5.4	3.2	4.1	3.8	4.3	-22.4	-21.6	-82	-80.9	19	19.7
2023/7/14	3.8	4.8	3.3	4.1	3.9	4.4	-22.5	-21.9	-82	-81	19	19.5
2023/7/15	3.9	4.7	3.3	4.1	3.9	4.3	-22.5	-22.2	-82	-80.9	19	19.2
2023/7/16	3.9	4.7	3.3	4.1	3.9	4.3	-22.5	-22.2	-82	-80.9	18.9	19
2023/7/17	3.9	5.1	3.1	4.1	3.9	4.3	-22.5	-20.9	-82	-80.9	18.8	19.4
2023/7/18	4.2	5.1	3.4	4.1	3.9	4.3	-21.2	-20.8	-82	-80.9	18.9	19.5
2023/7/19	4.3	5.1	3.4	4.3	3.9	4.4	-21.5	-20.9	-82	-79.7	19	19.5
2023/7/20	4.3	5.3	3.6	4.3	3.9	4.6	-21.7	-21.2	-81.5	-80.9	18.9	19.6
2023/7/21	4.3	6.7	3.5	4.2	4.3	5	-22.7	-21.4	-81.5	-80.6	18.9	19.7
2023/7/22	4.4	5.2	3.9	4.2	4.3	4.7	-22.7	-22.3	-81.5	-80.8	18.8	19.1
2023/7/23	4.4	5.3	3.9	4.2	4.3	4.6	-22.7	-22.3	-81.5	-80.5	18.8	18.8
2023/7/24	4.3	5.3	3.8	4.3	4.3	4.7	-22.7	-21.5	-81.5	-80.8	18.7	19.2
2023/7/25	4	5.3	3.9	4.2	4.1	4.7	-22.2	-21.5	-81.3	-80.6	18.8	19.3
2023/7/26	4	5.1	3.9	4.3	4.2	4.7	-22.4	-21.2	-81.5	-80.9	19	19.6
2023/7/27	3.9	5.1	4	4.6	4.3	4.8	-21.4	-21.2	-81.5	-80.9	19.2	19.9
2023/7/28	3.5	5.9	4.1	6.1	4.2	5.7	-23.9	-19.5	-81.5	-78.9	19.4	22.9
2023/7/29	3.2	4.8	4	5.4	4.2	4.6	-23.2	-22	-81.5	-80.9	20.4	23.9
2023/7/30	3.8	4.8	4	4.4	4.2	4.5	-22.7	-22.2	-81.5	-80.3	19.2	20.3
2023/7/31	3.9	5.1	4	4.9	4.2	4.7	-22.6	-19	-81.5	-79.4	19	19.6

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師:

藥學部 邱保祥 112/801
藥學部 邱予芹 112/801

總藥師:

藥學部 廖志峰 112/801

業管主管:

藥學部 周千澄 112/801
科主任