

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 112年 08 月

	一號冰箱		二號冰箱		三號冰箱		負20度冰箱		負80度冰箱		室溫1	
序號	0520-362		0520-292		1504-0305		1405-0193		1215-0006		1213-0007	
備用	0850-0835		0520-293		1504-0339		1213-0010		1330-0064		1405-0194	
範圍	2~8℃		2~8℃		2~8℃		負15~負25℃		≤負70℃		15~25℃	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2023/8/1	3.7	5.3	3.8	4.3	4.4	4.7	-20.9	-20.6	-81.5	-80.5	19.3	19.7
2023/8/2	3.7	5.1	3.9	4.7	4.4	5.1	-21.1	-20.7	-81.8	-80.7	19.2	19.6
2023/8/3	4.1	4.9	3.9	4.2	4.5	4.8	-21.1	-20.8	-81.8	-80.8	19.1	19.3
2023/8/4	3.9	4.9	3.9	6.3	4	4.9	-21.2	-20.6	-81.5	-80.7	19	19.5
2023/8/5	4.2	4.9	4.1	4.4	4	4.4	-21.2	-20.9	-81.8	-80.8	18.9	19.1
2023/8/6	4.1	4.9	4.1	4.3	4	4.4	-21.4	-20.9	-81.9	-80.6	18.8	18.9
2023/8/7	4.2	5.3	4.1	4.5	4	4.5	-21.4	-20.7	-81.9	-80.8	18.7	19.3
2023/8/8	4.3	5.1	4.1	4.4	4.1	4.5	-21.2	-20.7	-81.8	-80.8	19	19.6
2023/8/9	4.4	5.6	4.2	4.6	4.2	5.1	-21.2	-20.5	-81.7	-80.9	19.1	19.7
2023/8/10	3.8	5.2	4.3	4.7	4.6	5.1	-21.2	-20.5	-81.6	-80.7	19.2	19.8
2023/8/11	3.9	5	4.3	4.7	4.3	5.3	-21.2	-20.7	-81.7	-77.7	19.2	19.9
2023/8/12	4.2	5	4.4	4.7	4.3	4.6	-21.2	-20.9	-81.2	-80.6	19.2	19.7
2023/8/13	4.2	5	4.4	4.7	4.3	4.6	-21.2	-20.9	-81.4	-80.6	19.1	19.3
2023/8/14	4.1	5.1	4.4	4.9	4.3	5.2	-21.4	-20.9	-81.3	-80.8	18.6	19.4
2023/8/15	4	5.2	4.7	5.6	4.3	4.6	-21.2	-20.8	-81.5	-80.7	18.9	20.4
2023/8/16	4.2	5.3	4.8	5.4	4.3	4.9	-21.2	-20.7	-81.4	-79.4	19	19.7
2023/8/17	4.4	5.6	4.6	5.7	4.3	4.7	-21.2	-19.8	-81.2	-80.5	19.2	20.1
2023/8/18	5	5.6	4.5	5.3	4	4.9	-20.8	-19.8	-81.5	-79.9	17.8	19.6
2023/8/19	5.1	5.6	4.5	4.9	4	4.4	-20.6	-20.2	-81.6	-80.8	18.8	19
2023/8/20	5.1	5.6	4.5	4.8	4	4.4	-20.6	-20.2	-81.7	-80.8	18.6	18.8
2023/8/21	5	6	4.5	4.9	3.9	4.4	-20.7	-20.3	-81.7	-76.6	18.4	19
2023/8/22	4.9	5.8	3.4	4.9	4.1	4.5	-20.7	-20.2	-80.9	-80.7	18.6	18.9
2023/8/23	4.5	5.7	3.4	4.3	4.1	4.6	-20.8	-20.4	-80.9	-80.3	18.6	19.5
2023/8/24	4.5	5.4	3.2	4.2	4.2	4.9	-20.8	-19.8	-80.9	-79.8	18.7	19.9
2023/8/25	4.6	5.3	3.1	3.8	4.1	4.4	-20.7	-20.3	-80.7	-80.3	19.3	19.8
2023/8/26	4.6	5.3	3.3	3.8	4.1	4.4	-20.7	-20.3	-80.7	-80.1	19.2	19.4
2023/8/27	4.6	5.2	3.3	3.8	4.2	4.4	-20.7	-20.4	-80.8	-80.3	19.1	19.2
2023/8/28	4.6	5.4	3.3	3.8	4.2	4.5	-20.7	-20.1	-80.8	-80.6	19	19.6
2023/8/29	4.7	5.5	3.3	4.3	4.2	4.4	-20.6	-19.6	-80.7	-80.5	19.1	19.7
2023/8/30	4.8	5.7	3.7	4.3	4.1	4.6	-20.2	-19.9	-80.8	-80.6	19.3	19.8
2023/8/31	4.7	5.8	3.4	4.3	4.2	4.6	-20.2	-19.6	-80.7	-80.4	19.4	19.8

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師: 藥學部 邱保祥 112/09/01
藥學部 邱予芹 112/09/01

總藥師: 藥學部 廖志峰 112/09/01

業管主管:

藥學部 何沁沁 112/09/01