

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 112 年 08 月

	零號冰箱		四號冰箱		五號冰箱		八號冰箱		九號冰箱		室溫2	
序號	2052-0128		1611-0062		1715-0118		2123-0024		2123-0028		1715-0116	
備用	2052-0129		1611-0052		1715-0119		2123-0025		2123-0029		1715-0117	
範圍	2~8℃		2~8℃		2~8℃		2~8℃		2~8℃		15~25℃	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2023/8/1	3.9	4.4	4.2	4.9	3.8	5.5	5.1	5.9	5.8	6.2	21.5	21.9
2023/8/2	3.9	4.3	4.1	5.4	3.9	5.3	4.9	5.7	5.8	6.4	21.5	21.8
2023/8/3	3.9	4.3	4.1	4.7	4.7	5.3	5.3	5.7	5.7	6	21.4	21.7
2023/8/4	3.9	4.4	3.9	4.7	4.6	5.4	5.3	5.8	5.7	7.1	21.3	21.4
2023/8/5	4	4.4	3.9	4.6	4.6	5.3	5.5	5.8	5.9	6.1	21.3	21.4
2023/8/6	4	4.4	3.9	4.6	4.6	5.3	5.6	5.8	5.8	6.1	21.1	21.3
2023/8/7	4	4.4	3.9	4.6	4.4	5.3	5.3	5.9	4.6	6	21.1	21.5
2023/8/8	3.8	4.4	3.8	4.6	4.4	5.6	5.3	5.7	4.5	5.6	21.4	21.8
2023/8/9	3.8	4.5	3.9	4.6	4.2	5.5	5.2	5.8	4.5	5.2	21.7	22
2023/8/10	3.8	4.3	3.9	4.6	4.2	5.2	4.7	5.7	4.8	5.5	21.9	22.1
2023/8/11	3.8	4.3	3.9	5.1	4.2	5.3	4.7	5.3	5	5.5	21.9	22.1
2023/8/12	3.8	4.3	3.9	4.6	4.2	5.1	4.6	5.3	4.9	5.4	21.7	22
2023/8/13	3.8	4.3	3.9	4.7	4.2	5.1	4.6	5.2	4.9	5.4	21.5	21.8
2023/8/14	3.7	4.3	3.9	4.7	4.2	5.6	4.6	5.2	4.9	5.6	21.3	21.5
2023/8/15	3.7	4.5	3.9	4.9	4.6	6.1	4.6	5.3	4.9	5.5	21.3	21.7
2023/8/16	3.7	4.5	3.7	4.7	4.9	5.9	4.6	5.5	4.9	5.5	21.7	21.8
2023/8/17	4.2	4.6	3.7	4.7	4.8	5.8	4.8	5.6	4.7	5.2	21.7	21.9
2023/8/18	3.7	4.6	4	4.7	4.7	5.8	4.7	5.4	4.6	5.2	21.3	21.7
2023/8/19	3.7	4.2	4	4.7	4.7	5.6	4.7	5.3	4.5	5	20.9	21.3
2023/8/20	3.7	4.2	3.9	4.7	4.7	5.6	4.7	5.3	4.5	5	20.7	20.9
2023/8/21	3.6	4.3	4	4.7	4.6	5.7	4.7	5.3	4.5	5	20.7	21.2
2023/8/22	3.6	4.2	4.1	4.9	4.6	5.5	4.7	5.3	4.5	5.1	20.9	21.3
2023/8/23	3.6	4.2	4	4.8	4.6	5.5	4.7	5.3	4.6	5.2	21.1	21.8
2023/8/24	3.6	5	4	4.7	4.3	5.7	4.7	5.3	4.7	5.3	21.5	21.8
2023/8/25	4.6	5.3	3.9	4.7	4.2	5.3	4.7	5.3	4.6	5.6	21.5	21.9
2023/8/26	4.6	4.9	3.9	4.6	4.4	5.3	4.7	5.3	4.5	4.9	21.3	21.7
2023/8/27	4.5	4.9	3.9	4.6	4.4	5.3	4.7	5.3	4.5	4.9	21.2	21.3
2023/8/28	4.5	4.9	3.9	4.7	4.3	5.3	4.7	5.3	4.5	5.1	21.1	21.4
2023/8/29	4.5	5.8	3.9	4.6	4.4	5.7	4.4	5.3	4.5	5.1	21.1	21.7
2023/8/30	5.6	6.6	3.9	4.6	4.2	5.4	4.4	5	4.4	5.2	21.4	21.7
2023/8/31	6.2	6.6	3.9	4.6	4.1	5.4	4.4	5.2	4.4	5.2	21.4	21.7

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師:

溫度均在正常範圍內

藥學部 邱保祥
師(三)級藥師
藥學部 邱予芹
師(三)級藥師

總藥師:

藥學部 廖志峰
師(二)級藥師

業管主管:

藥學部 何沁沁
師(一)級藥師