

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 112年 09月

序號	一號冰箱		二號冰箱		三號冰箱		負20度冰箱		負80度冰箱		室溫1	
	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
備用	0850-0835		0520-293		1504-0339		1213-0010		1330-0064		1405-0194	
範圍	2~8°C		2~8°C		2~8°C		負15~負25°C		≤負70°C		15~25°C	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2023/9/1	4.8	5.6	3.6	4.2	4	4.4	-19.6	-18.3	-80.7	-80.4	19.5	20.8
2023/9/2	4.8	5.5	3.6	4.2	4	4.4	-19.6	-19.4	-80.7	-80.3	19.4	19.5
2023/9/3	4.8	5.4	3.6	4.2	4	4.4	-19.6	-19.6	-80.7	-80.3	19	19.4
2023/9/4	4.8	5.7	3.6	4.3	4	4.7	-19.7	-19.1	-80.8	-80.3	19.4	19.9
2023/9/5	5.1	5.9	3.5	4.2	4.1	4.6	-19.7	-19.1	-80.9	-79.8	19.4	19.9
2023/9/6	5	5.7	3.4	4.2	4	4.8	-19.8	-18.9	-80.9	-80.2	19.4	19.8
2023/9/7	4.8	5.8	3.1	4.8	4	4.4	-19.8	-19	-80.9	-80.3	19.3	19.9
2023/9/8	4.8	6.6	2.9	4.6	4	5	-19.7	-18.8	-80.9	-80.4	19.3	19.9
2023/9/9	5.1	5.6	4.1	4.4	4.1	4.4	-19.3	-19.1	-80.9	-80.2	19.2	19.4
2023/9/10	5.1	5.6	4.2	4.4	4	4.4	-19.3	-19.2	-80.9	-80.3	19	19.2
2023/9/11	5.1	5.7	4.2	4.5	4	4.4	-20.3	-18.8	-80.8	-80	18.9	19.4
2023/9/12	4.3	6	3.9	4.8	4.1	4.5	-19.9	-18.4	-81.3	-74.8	19	19.6
2023/9/13	5.3	6.1	3.9	4.4	4.1	4.4	-19.6	-18.6	-81.5	-80.8	19.1	19.4
2023/9/14	5.3	6.1	3.9	4.4	4.1	4.6	-19.6	-16.3	-81.5	-80.9	19.1	19.6
2023/9/15	5.3	6	3.9	4.7	4.2	4.8	-20.6	-18.7	-81.5	-80.9	19.2	19.9
2023/9/16	5.3	5.8	3.9	4.3	4.3	4.7	-20.7	-20.4	-81.5	-80.8	19	19.2
2023/9/17	5.2	5.8	3.9	4.3	4.4	4.6	-20.7	-20.4	-81.2	-80.7	18.9	19
2023/9/18	5.2	5.9	3.9	4.3	4.4	4.6	-20.9	-20.4	-81.4	-80.5	18.8	19.6
2023/9/19	5.2	6.1	3.9	4.3	3.9	4.6	-21.1	-20.7	-81.4	-80	19.1	19.8
2023/9/20	4.9	5.8	3.9	4.7	3.9	4.4	-21.1	-19.9	-81.2	-81	19.2	19.7
2023/9/21	5.2	5.9	4.2	4.7	4	4.7	-20.7	-19.7	-81.2	-80.9	19.2	19.8
2023/9/22	4.5	6	4.2	4.5	4.2	4.5	-20.7	-20.2	-81.2	-80.4	19.2	20.2
2023/9/23	4.5	5.2	4.2	4.6	4.2	4.6	-20.4	-19.1	-82	-74.1	19.2	19.8
2023/9/24	4.7	5.2	4.2	4.4	4.2	4.5	-20.2	-19.9	-81.9	-80.1	19.3	19.4
2023/9/25	4.6	5.2	4.1	4.5	4.2	4.5	-20.4	-19.9	-82	-80	19.3	19.9
2023/9/26	4.5	5.3	4	4.5	4.2	4.8	-20.4	-20.2	-82	-79.8	19.3	20.2
2023/9/27	3.5	5.4	4.1	4.6	4.2	4.6	-20.4	-19.8	-81.9	-79.5	19.4	20.1
2023/9/28	3.4	4.7	4.1	4.6	3.8	4.8	-20.4	-19.9	-81.9	-79.9	19.5	20.1
2023/9/29	3.9	4.3	4.2	4.4	4.1	4.5	-20.4	-20.2	-81.9	-78.7	19.2	19.6
2023/9/30	3.9	4.3	4.2	4.4	4.2	4.5	-20.4	-20.2	-81.9	-79.9	19.2	19.3

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師: 藥學部 邱保祥 112 總藥師: 藥學部 廖志峰 112
藥學部 邱子芹 112 藥學部 何沁沁 112
藥學部 張豫立 112