

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 112年 10月

	一號冰箱		二號冰箱		三號冰箱		負20度冰箱		負80度冰箱		室溫1	
序號	0520-362		0520-292		1504-0305		1405-0193		1215-0006		1213-0007	
備用	0850-0835		0520-293		1504-0339		1213-0010		1330-0064		1405-0194	
範圍	2~8℃		2~8℃		2~8℃		負15~負25℃		≤負70℃		15~25℃	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2023/10/1	3.9	4.3	4.2	4.4	4.1	4.5	-20.4	-20.2	-81.8	-78.5	19.2	19.2
2023/10/2	3.6	5.6	4.1	4.5	4.2	4.6	-21.2	-18.8	-81.8	-78.1	19.2	19.8
2023/10/3	3.3	5.9	3.2	4.8	4.2	4.6	-21.2	-19.9	-81.4	-79.4	19.2	19.8
2023/10/4	3.3	3.9	3.2	4.2	4.2	4.8	-20.2	-19.3	-80.6	-78.9	19.3	19.9
2023/10/5	2.4	4	3.1	5.4	4.2	4.7	-20.2	-19.9	-80.7	-78.3	19.4	20.8
2023/10/6	3.1	3.9	3.1	4	4.2	4.7	-20.2	-19.3	-80.7	-79.6	19.6	20.5
2023/10/7	3.3	3.7	3.2	4	4.1	4.5	-19.9	-19.8	-80.7	-77.9	19.3	19.7
2023/10/8	3.3	3.7	3.2	4	4.1	4.5	-19.9	-19.9	-80.6	-78.6	19.1	19.5
2023/10/9	3.3	3.7	3.2	4	4.1	4.5	-19.9	-19.9	-80.5	-79.5	19.4	19.5
2023/10/10	3.3	3.7	3.2	4	4.1	4.5	-19.9	-19.9	-80.7	-79.4	19.4	19.4
2023/10/11	3.3	4.2	3.2	4.1	4.2	4.7	-20.7	-19.9	-80.9	-77.9	19.2	20.1
2023/10/12	3.3	5.3	2.7	5.9	4.1	4.7	-20.7	-19.9	-80.6	-79.4	19.3	20
2023/10/13	3.6	4.3	2.8	3.7	3.9	4.6	-20.2	-19.6	-80.7	-78.2	19.4	20.2
2023/10/14	3.6	4.3	2.8	3.7	3.9	4.4	-20.2	-19.9	-80.6	-79.4	19.3	19.6
2023/10/15	3.7	4.3	2.8	3.7	3.9	4.4	-20.2	-20	-80.7	-79.4	18.9	19.3
2023/10/16	3.3	4.9	2.8	3.7	3.9	4.4	-20.3	-19.9	-80.7	-79.3	18.8	19.3
2023/10/17	3.4	4.5	2.8	3.7	3.9	4.9	-20.4	-19.7	-80.7	-78.9	18.4	19.4
2023/10/18	3.9	4.6	2.8	5.6	3.9	4.5	-20.4	-19	-80.7	-79.4	19.1	20.1
2023/10/19	3.9	4.7	5.1	6	4	4.5	-20.2	-19	-80.7	-79.4	19.4	20.3
2023/10/20	3.9	4.6	4.8	5.5	4.1	4.9	-19.6	-19	-80.7	-79	19.7	20.4
2023/10/21	3.9	4.4	5.1	5.4	4.2	4.6	-19.6	-19.3	-80.7	-78.4	18.7	19.8
2023/10/22	3.9	4.4	5.1	5.3	4.2	4.6	-19.6	-19.3	-80.7	-79.4	18.4	19.4
2023/10/23	3	4.7	5.1	5.5	4.2	4.7	-20.6	-19.4	-80.6	-79.4	18.6	19.7
2023/10/24	2.9	4	4.8	5.6	4.2	4.6	-20.4	-19.3	-80.7	-78.1	19	19.9
2023/10/25	2.9	4.8	4.8	6.1	4.1	4.8	-20.2	-19.4	-80.6	-79.3	19	19.8
2023/10/26	4.3	4.8	5.3	6.2	4.1	4.4	-20.4	-19.4	-80.6	-79.3	18.9	19.8
2023/10/27	4.4	4.8	5.7	6.2	4.1	4.6	-20.4	-19.9	-80.7	-78.3	18.9	19.6
2023/10/28	4.4	4.8	5.6	5.8	4.1	4.5	-20.4	-20.2	-80.6	-79.3	18.3	18.9
2023/10/29	4.4	4.8	5.6	5.8	4.1	4.4	-20.4	-20.2	-80.7	-79.4	18.1	18.3
2023/10/30	4.4	4.9	5.6	5.9	4.1	4.6	-20.4	-19.1	-80.7	-79.4	17.9	19.2
2023/10/31	4.3	4.8	5.6	5.9	4.2	4.6	-19.5	-18.9	-80.7	-79.4	18.4	19.2

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師: 藥學部 邱保祥 (112)
師(三)級藥師 邱保祥 (101)
藥學部 邱予芹 (112)
師(三)級藥師 邱予芹 (101)

總藥師: 藥學部 廖志峰 (112)
師(二)級藥師 廖志峰 (102)
1340

業管主管: 藥學部 何沁沁 (106)
師(一)級藥師 何沁沁 (0830)

藥學部 張豫立
主任 張豫立