

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 112年 11月

	一號冰箱		二號冰箱		三號冰箱		負20度冰箱		負80度冰箱		室溫1	
序號	0520-362		0520-292		1504-0305		1405-0193		1215-0006		1213-0007	
備用	0850-0835		0520-293		1504-0339		1213-0010		1330-0064		1405-0194	
範圍	2~8℃		2~8℃		2~8℃		負15~負25℃		≤負70℃		15~25℃	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2023/11/1	3.9	4.4	5.6	5.9	4.3	4.9	-20.7	-18.2	-79.8	-78.9	19.5	19.8
2023/11/2	3.8	4.4	5.6	6.1	4.3	4.8	-20.8	-19.9	-79.8	-79	19.5	19.8
2023/11/3	3.9	4.5	5.8	6.2	4.4	4.8	-20.8	-20.1	-79.8	-77.8	19.4	19.7
2023/11/4	4.1	4.5	5.8	5.9	4.4	4.7	-20.8	-20.5	-79.8	-77.7	19.4	19.6
2023/11/5	4.1	4.5	5.8	5.9	4.4	4.7	-20.7	-20.5	-79.8	-78.8	19.2	19.4
2023/11/6	3.9	4.7	5.9	6.1	4.3	4.8	-20.7	-19.9	-79.8	-77.7	19.3	19.7
2023/11/7	4	4.6	5.7	6.2	4.2	4.9	-20.7	-19.9	-79.8	-78.4	19.6	19.9
2023/11/8	3.8	4.5	5.7	6.1	4.3	4.9	-20.7	-19.7	-79.8	-78	19.8	20.1
2023/11/9	3.1	5.6	5.8	6.1	4.4	5.1	-20.7	-19.8	-79.8	-78.1	19.8	20.1
2023/11/10	3.1	4.2	5.8	6.4	4.3	5.1	-20.4	-19.8	-79.8	-78.2	19.8	20.1
2023/11/11	3.1	4.1	6	6.1	4.3	4.7	-20.2	-19.9	-79.8	-78.1	19.9	20.1
2023/11/12	2.9	4.1	6	6.2	4.3	4.7	-20.2	-19.9	-79.8	-78.8	19.9	20.1
2023/11/13	3.1	4.2	5.9	6.1	4.3	4.8	-20.7	-19.9	-79.8	-78.3	19.7	19.9
2023/11/14	2.7	4.2	5.9	6.2	4.3	5.4	-20.7	-19.6	-79.8	-78.3	19.7	19.9
2023/11/15	2.7	4.3	6	6.6	4.2	4.8	-20.6	-19.1	-79.8	-78	19.7	20.1
2023/11/16	3.6	4.6	4.8	6.4	4.1	5.7	-20.7	-20.2	-79.8	-78.4	19.9	20.2
2023/11/17	2.8	5.2	4.9	5.4	4.1	4.6	-20.9	-20.4	-79.9	-77.9	19.7	20.1
2023/11/18	3.1	4.1	4.9	5.3	4.2	4.5	-20.9	-20.7	-79.8	-77.9	19.4	19.7
2023/11/19	3	4.1	4.9	5.4	4.2	4.5	-20.9	-20.7	-79.6	-77.9	19.4	19.5
2023/11/20	3	4.4	4.9	5.7	4.2	4.7	-20.9	-19.9	-79.8	-77.5	19.5	19.9
2023/11/21	3.9	4.6	5.1	6	4.3	4.8	-20.7	-19.9	-79.8	-78	19.8	20.2
2023/11/22	3.9	4.6	5.8	6.5	4.3	4.9	-20.4	-19.8	-79.8	-77.8	20.1	20.3
2023/11/23	3.9	4.6	5.6	6.2	4.2	4.6	-20.3	-19.7	-79.8	-77.9	20.1	20.3
2023/11/24	3.9	4.7	3.7	5.7	4.2	4.7	-20.4	-19.7	-79.8	-78	19.8	20.1
2023/11/25	3.9	4.4	3.7	4.5	4.2	4.5	-20.6	-20.3	-79.8	-78.9	19.5	19.7
2023/11/26	3.9	4.4	3.7	4.5	4.2	4.5	-20.6	-20.4	-79.8	-77.7	19.4	19.5
2023/11/27	3.9	4.7	3.7	4.5	3.6	4.8	-21.5	-19.6	-79.9	-77.7	19.4	19.6
2023/11/28	3.9	4.6	3.7	4.7	3.4	4.6	-20.9	-19.9	-79.9	-77.2	19.4	19.6
2023/11/29	3.9	4.6	3.5	4.2	3.6	4.3	-20.7	-19.3	-77.9	-77	19.5	19.7
2023/11/30	3.9	4.4	3.5	4.3	3.6	4.2	-19.6	-19.1	-78.1	-77	19.4	19.6

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師:

藥學部 邱保祥 112/201
藥學部 邱子芹 112/201

總藥師:

藥學部 廖志峰 112/205

業管主管:

藥學部 何沁沁 112/206

藥學部 周千澐 112/207