

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 112年 12月

	一號冰箱		二號冰箱		三號冰箱		負20度冰箱		負80度冰箱		室溫1	
序號	0520-362		0520-292		1504-0305		1405-0193		1215-0006		1213-0007	
備用	0850-0835		0520-293		1504-0339		1213-0010		1330-0064		1405-0194	
範圍	2~8℃		2~8℃		2~8℃		負15~負25℃		≤負70℃		15~25℃	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2023/12/1	3.8	4.4	3.7	4.3	3.6	4.1	-18.7	-18.1	-78.1	-77	19.3	19.5
2023/12/2	3.8	4.4	3.7	4.2	3.6	4.2	-18.7	-18.6	-78.1	-76.4	18.9	19.3
2023/12/3	3.8	4.4	3.7	4.2	3.6	4.2	-18.7	-18.7	-78.1	-76.2	18.8	18.9
2023/12/4	3.8	4.4	3.7	4.3	3.5	5.1	-19.4	-18.7	-78.1	-76.7	18.8	19.2
2023/12/5	3.6	5.5	3.4	4.5	3.4	4.2	-20.6	-19.3	-77.9	-77.2	19	19.3
2023/12/6	3.6	4.7	3.9	4.5	3.5	4.3	-20.6	-19.8	-77.9	-77.2	19.2	19.4
2023/12/7	3.7	4.6	4	4.6	3.4	4.2	-20.2	-19.4	-78	-77.2	19.2	19.4
2023/12/8	3.7	4.8	4.1	4.7	3.4	4.2	-20.4	-19.9	-78.1	-77.2	19.2	20.1
2023/12/9	4.2	4.7	4	4.4	3.6	4.3	-20.4	-20.3	-78.1	-77.2	19.2	19.3
2023/12/10	4.2	4.7	4	4.4	3.6	4.2	-20.4	-20.2	-78.1	-76.9	19.2	19.4
2023/12/11	4.2	5.1	3.3	4.6	3.6	4.4	-20.4	-19.1	-78.1	-75.9	19.4	19.7
2023/12/12	3.9	5.2	3.1	3.9	3.6	4.3	-20.3	-18.9	-77.9	-76.9	19.7	19.9
2023/12/13	4	4.7	3.1	4.1	3.1	4.7	-20.2	-19.9	-77.9	-76.5	19.8	19.9
2023/12/14	4.1	4.7	3.1	4.3	3.3	4.3	-20.2	-19.3	-77.9	-76.6	19.8	20.1
2023/12/15	4.2	4.7	3.2	4.3	3.6	4.5	-19.7	-19.1	-77.9	-76.9	19.7	19.9
2023/12/16	4.1	4.7	3.5	4	3.8	4.3	-19.7	-19.6	-78.1	-75.9	19.3	19.7
2023/12/17	4.1	4.7	3.5	4	3.9	4.3	-19.8	-19.6	-78.1	-76.2	18.9	19.3
2023/12/18	3.9	4.9	3.5	4.3	3.8	4.8	-19.9	-19.6	-78.1	-76.2	19	19.6
2023/12/19	3.8	4.7	3.7	4.6	4	4.6	-19.9	-19.4	-78.1	-77.2	19.5	20.1
2023/12/20	3.8	4.6	3.9	4.6	4	4.6	-19.9	-19.6	-78.2	-76.8	19.6	19.9
2023/12/21	3.9	4.6	3.8	4.5	4	4.6	-20.1	-19.6	-78.2	-76.2	19	19.6
2023/12/22	4.1	5.1	3.9	4.5	4	4.6	-20.2	-19.6	-78.1	-77.1	18.5	19
2023/12/23	4.2	5.7	3.9	5.3	4.1	5.6	-20.2	-19.9	-78.2	-73.8	18.4	19
2023/12/24	4.2	4.7	3.9	4.5	4.2	4.6	-20.2	-19.9	-78	-76.7	18.9	19.2
2023/12/25	3.9	4.7	3.9	4.5	4.1	4.6	-20.2	-19.8	-78	-77.2	18.7	19
2023/12/26	3.9	4.7	3.9	4.5	4.1	4.8	-19.9	-19.8	-77.9	-77.2	18.9	19.6
2023/12/27	3.9	5	3.9	4.6	4.2	4.7	-19.9	-19.4	-77.9	-77.2	19.6	20.1
2023/12/28	4.1	4.7	3.9	4.7	4.2	4.7	-19.9	-19.4	-78.1	-77.2	20.1	20.4
2023/12/29	4.1	4.7	3.9	4.7	4.1	4.7	-19.9	-19.6	-78.2	-77.2	20.3	20.6
2023/12/30	4.2	4.7	3.9	4.6	4	4.5	-19.9	-19.4	-78.1	-77.2	20.5	22
2023/12/31	4.2	4.7	3.9	4.5	4.1	4.5	-19.9	-19.7	-78.1	-75.9	20.6	21.3

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師: 藥學部 邱保祥
藥學部 邱子芹

總藥師: 藥學部 廖志峰

業管主管:

藥學部 何沁沁

藥學部 周千澄