

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 112 年 12 月

	零號冰箱		四號冰箱		五號冰箱		八號冰箱		九號冰箱		室溫2	
序號	2052-0128		1611-0062		1715-0118		2123-0024		2123-0028		1715-0116	
備用	2052-0129		1611-0052		1715-0119		2123-0025		2123-0029		1715-0117	
範圍	2~8℃		2~8℃		2~8℃		2~8℃		2~8℃		15~25℃	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2023/12/1	4.5	4.9	4.2	4.9	3.9	4.5	5.4	6.3	3.8	4.5	19.4	19.9
2023/12/2	4.5	4.8	4.2	4.7	3.9	4.5	5.5	5.8	3.9	4.5	19.4	19.7
2023/12/3	4.5	4.9	4.2	4.8	3.9	4.5	5.4	5.8	3.9	4.6	19.2	19.4
2023/12/4	4.5	5.2	4.2	4.9	3.9	4.5	5.4	5.9	3.8	4.6	19.2	19.6
2023/12/5	4.9	5.6	4.4	5.2	3.8	4.4	5.4	6.1	3.5	4.9	19.4	19.7
2023/12/6	4.9	5.6	4.4	5.3	3.9	4.5	5.5	5.8	4.2	5.3	19.5	19.8
2023/12/7	4.9	5.3	4.4	5.1	3.9	4.6	5.3	5.7	4.2	4.8	19.5	19.7
2023/12/8	4.9	5.7	4.4	5.2	3.9	4.4	5.4	6	4.2	4.8	19.4	19.8
2023/12/9	4.9	5.2	4.4	5.1	3.9	4.4	5.7	5.9	4.2	4.8	19.6	19.7
2023/12/10	4.9	5.2	4.4	5.1	3.9	4.4	5.7	5.9	4.2	4.8	19.5	19.8
2023/12/11	4.8	5.2	4.3	5.3	3.9	4.5	5.8	6.8	4.2	5.1	19.8	20.2
2023/12/12	4.6	5.2	4.4	5.2	3.9	4.6	6.4	6.7	4.2	4.9	20.1	20.4
2023/12/13	4.4	4.9	4.4	5.2	3.9	4.6	6.4	6.7	4.2	5.1	20.2	20.5
2023/12/14	4.4	4.9	4.5	5.3	4.1	4.8	6.5	6.8	4.3	5.2	20.3	20.6
2023/12/15	4.7	5.3	4.5	5.3	4.2	4.6	6.2	6.7	4.3	5.3	20.3	20.5
2023/12/16	4.9	5.1	4.5	5.2	4.1	4.5	6.2	6.4	4.3	5.1	19.7	20.3
2023/12/17	4.9	5.1	4.5	5.2	4.2	4.5	6.1	6.3	4.4	5.1	19.6	19.7
2023/12/18	4.9	5.2	4.6	5.3	4.1	4.8	6.1	6.5	4.3	5.3	19.6	20.2
2023/12/19	4.8	5.2	4.6	5.4	4.3	5.1	5.5	6.4	4.2	5	20.2	20.7
2023/12/20	4.9	6.1	4.7	5.4	4.3	4.9	5.6	6	4.2	5.1	20.3	20.6
2023/12/21	5.6	5.9	4.7	5.3	4	4.8	5.4	6	4.2	5.1	19.6	20.3
2023/12/22	5.4	5.8	4.7	5.6	4.1	4.5	5.4	5.9	4.3	4.9	19	19.6
2023/12/23	5.4	6.7	4.7	5.9	4	5.8	5.6	7.2	4.4	6	19	19.5
2023/12/24	5.5	5.8	4.7	5.3	4	4.6	5.6	6.1	4.4	5.6	19.4	19.6
2023/12/25	5.4	5.9	4.6	5.4	4.1	4.6	3.5	6.6	3.9	4.9	19.2	19.6
2023/12/26	5.8	6.3	4.6	5.2	4.1	4.5	3.7	4.8	4	5.4	19.5	20.2
2023/12/27	5.8	6.2	4.4	5.2	4	4.5	3.3	4.7	4.1	4.9	20.2	20.6
2023/12/28	5.8	6.1	4.4	5.1	4.1	4.5	3.3	4.7	4.1	4.9	20.5	20.8
2023/12/29	5.9	6.1	4.4	5.3	4.1	4.7	3.4	4.8	4.1	4.9	20.7	20.9
2023/12/30	5.9	6.3	4.5	5.2	3.9	4.5	3.2	4.7	4.2	5	20.7	22.3
2023/12/31	5.9	6	4.6	5.2	4.1	4.5	3.5	4.7	4.2	4.9	20.6	21.4

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

溫度均在正常範圍內

承辦藥師:

藥學部 邱保祥
師(三)級藥師

總藥師

藥學部 廖志峰
師(一)級藥師

業管主管:

藥學部 何沁沁
師(一)級藥師

藥學部 周千澄
主任