

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 113 年 01 月

	零號冰箱		四號冰箱		五號冰箱		八號冰箱		九號冰箱		室溫2	
序號	2052-0128		1611-0062		1715-0118		2123-0024		2123-0028		1715-0116	
備用	2052-0129		1611-0052		1715-0119		2123-0025		2123-0029		1715-0117	
範圍	2~8℃		2~8℃		2~8℃		2~8℃		2~8℃		15~25℃	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2024/1/1	5.8	5.9	4.7	5.2	4.1	4.5	3.6	4.7	4.2	4.9	20.1	20.6
2024/1/2	5.8	6	4.5	5.2	4.2	4.6	3.6	4.7	5.3	5.9	20.1	20.7
2024/1/3	5.8	6.1	4.5	5.3	4.2	4.6	3.5	4.7	5.3	5.8	20.2	20.5
2024/1/4	5.7	6.1	4.5	5.3	3.9	4.7	3.6	4.8	5.3	5.9	19.9	20.4
2024/1/5	5.8	6.4	4.4	5.2	3.9	4.9	3.6	4.8	5.3	6	20.1	20.5
2024/1/6	5.9	6	4.4	5.1	4.2	4.5	3.6	4.8	5.7	6	20.3	20.5
2024/1/7	5.8	5.9	4.4	5.1	4.2	4.5	3.7	4.9	5.6	5.9	19.9	20.3
2024/1/8	5.7	6.1	4.4	5.2	3.9	4.6	3.7	4.8	5.6	6.2	19.9	20.3
2024/1/9	5.7	6.6	3.7	6.2	4.1	4.7	3.2	4.8	5.7	6.3	20.2	20.6
2024/1/10	5.7	6.1	3.4	4.5	4.1	4.8	3.5	4.9	4.4	5.9	20.2	20.5
2024/1/11	5.7	5.9	3.4	4.6	3.8	4.7	3.7	4.9	4.4	5.2	19.9	20.2
2024/1/12	5.7	6.1	3.3	4.5	3.9	5.1	2.9	4.9	4.4	5.3	19.9	20.3
2024/1/13	5.7	5.8	3.3	4.4	4	4.6	3.4	4.7	4.6	5.3	20.1	20.3
2024/1/14	5.7	5.8	3.4	4.5	4.1	4.6	3.4	4.7	4.5	5.2	19.9	20.1
2024/1/15	4.9	7.2	3.3	4.6	4	4.7	3.3	4.7	4.6	5.5	20.1	20.5
2024/1/16	6.2	6.7	3.4	4.6	4.1	4.8	3.3	4.8	4.7	5.8	20.2	20.6
2024/1/17	6.2	6.7	3.4	4.6	4.2	4.6	3.3	4.7	4.7	5.9	20.5	20.9
2024/1/18	6.2	6.4	3.5	4.5	4.2	4.6	3.3	4.8	4.7	5.3	20.6	21.1
2024/1/19	6.2	6.4	3.5	4.7	4.2	4.8	3.4	4.7	4.7	5.3	20.8	21.2
2024/1/20	6.2	6.3	3.7	4.7	4.3	4.7	3.5	4.8	4.7	5.3	20.9	21.1
2024/1/21	6.2	6.3	3.8	4.7	4.3	4.7	3.5	4.8	4.7	5.3	20.6	21
2024/1/22	6.1	6.3	3.8	4.8	4.3	4.9	3.5	4.9	4.7	5.3	20.3	20.6
2024/1/23	6.1	6.3	4.1	5.1	4.4	4.9	3.8	4.9	4.7	5.3	19.6	20.3
2024/1/24	5.8	6.5	4.2	5.1	4.3	5.3	3.9	4.9	4.6	5.1	19.2	19.6
2024/1/25	5.7	6.4	3.6	4.9	4.2	4.7	3.3	4.9	4.4	5.7	19.2	19.5
2024/1/26	5.7	6.1	3.3	4.9	4.2	4.8	3.3	4.9	4.4	4.9	19.3	19.6
2024/1/27	5.8	5.9	3.6	4.5	4.3	4.7	3.3	4.8	4.4	4.8	19.4	19.5
2024/1/28	5.7	5.9	3.6	4.5	4.3	4.7	3.3	4.8	4.4	4.8	19.2	19.5
2024/1/29	5.7	6.1	3.6	4.7	4.3	4.7	3.3	4.8	4.4	4.9	19.3	19.9
2024/1/30	5.9	6.2	3.6	4.7	4.2	4.9	3.4	4.8	4.4	5.4	19.9	20.5
2024/1/31	5.9	6.6	3.6	4.6	4.2	4.9	3.4	4.9	4.7	5.2	20.4	20.8

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師: 藥學部 邱保祥 (13/20)
藥學部 邱子芹 (13/20)

總藥師:

藥學部 廖志峰 (13/20)
1630

業管主管:

藥學部 何沁沁 (02/02)
1100 W