

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 113年 02 月

	一號冰箱		二號冰箱		三號冰箱		負20度冰箱		負80度冰箱		室溫1	
序號	0520-362		0520-292		1504-0305		1405-0193		1215-0006		1213-0007	
備用	0850-0835		0520-293		1504-0339		1213-0010		1330-0064		1405-0194	
範圍	2~8℃		2~8℃		2~8℃		負15~負25℃		≤負70℃		15~25℃	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2024/2/1	3.3	4.3	3.4	4.3	3.3	4.3	-20.4	-18.5	-80.7	-79.3	19.1	19.4
2024/2/2	3.2	4.6	3.4	4.3	3.3	4.2	-20.4	-19.3	-80.7	-79.3	18.9	19.4
2024/2/3	4	4.6	3.7	4.2	3.7	4.2	-19.7	-19.7	-80.7	-79.3	18.9	19.1
2024/2/4	4	4.6	3.7	4.2	3.7	4.2	-19.8	-19.6	-80.7	-79.3	18.8	18.9
2024/2/5	3.8	4.5	3.7	4.3	3.3	4.4	-20.2	-18.9	-80.7	-79.2	18.7	19.3
2024/2/6	3.8	4.5	3.7	4.3	3.3	4	-19.6	-19.1	-80.5	-79.3	18.6	19.1
2024/2/7	3.8	4.4	3.7	4.3	3.3	4	-19.6	-18.9	-80.6	-78	18.6	19.2
2024/2/8	3.8	4.4	3.9	4.3	3.3	3.9	-19.5	-19.4	-80.5	-79.2	18.4	18.7
2024/2/9	3.8	4.3	3.8	4.3	3.3	4	-19.6	-19.4	-80.7	-79.2	18.2	18.4
2024/2/10	3.8	4.4	3.8	4.3	3.3	4	-19.6	-19.4	-80.7	-77.9	18	18.2
2024/2/11	3.8	4.3	3.9	4.3	3.4	4	-19.6	-19.4	-80.7	-77.5	17.7	18
2024/2/12	3.8	4.3	3.9	4.3	3.4	4	-19.6	-19.4	-80.7	-78	17.7	17.8
2024/2/13	3.8	4.4	3.9	4.3	3.4	4	-19.6	-19.4	-80.7	-79	17.7	17.8
2024/2/14	3.8	4.6	3.7	4.3	3.3	4.1	-19.7	-19.1	-80.6	-79.2	17.8	18.9
2024/2/15	3.7	4.6	3.7	4.3	3.3	3.9	-20.8	-19.5	-80.5	-79.2	18.6	19.9
2024/2/16	3.6	4.4	3.7	4.2	3.4	4.2	-20.4	-18.6	-80.6	-79.2	18.4	19.1
2024/2/17	3.6	5.1	3.7	4.2	3.5	4.9	-19.9	-19.3	-80.7	-79.2	18.3	19.4
2024/2/18	3.8	4.3	3.7	4.2	3.5	4.1	-19.6	-19.4	-80.7	-78.7	18.8	19.4
2024/2/19	3.6	4.4	3.3	4.2	3.4	4	-19.7	-19.4	-80.7	-77.7	19.3	19.9
2024/2/20	3.3	4.8	3.3	4.1	3.4	4.2	-19.7	-18.8	-80.7	-77.7	19.9	20.4
2024/2/21	3.3	4.7	3.3	4.3	3.3	4.2	-19.6	-19.1	-80.7	-79.2	19.1	20.2
2024/2/22	3.7	4.4	3.3	4	3.3	5.2	-19.7	-18.3	-80.5	-79.2	18.7	19.2
2024/2/23	3.7	4.4	3.3	4	3.3	4.2	-19.6	-18.9	-80.6	-78.5	18.4	18.8
2024/2/24	3.9	4.4	3.6	4	3.3	4	-19.6	-19.4	-80.7	-77.9	18.2	18.4
2024/2/25	3.9	4.4	3.6	4	3.3	4.1	-19.6	-19.6	-80.7	-79.3	17.4	18.2
2024/2/26	3.7	4.7	3.7	4.3	3.3	4.1	-19.7	-19.5	-80.7	-79.3	17.3	18.2
2024/2/27	3.6	4.4	4	4.7	3.4	4.1	-19.7	-19.1	-80.7	-79	17.9	18.6
2024/2/28	3.6	4.2	4.4	4.7	3.4	4.1	-19.6	-19.4	-80.7	-79.3	18.3	18.7
2024/2/29	3.6	4.4	4.2	4.7	3.3	4.2	-20.2	-19.6	-80.7	-78.4	18.2	19.3

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師: 藥學部 邱保祥 (113) 301
 藥學部 邱子芹 (113) 304

總藥師:

藥學部 廖志峰 (113) 1420

業管主管:

藥學部 何沁沁 (113) 305

藥學部 周千澄 (113) 305