

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 113年 10月

	一號冰箱		二號冰箱		三號冰箱		負20度冰箱		負80度冰箱		室溫1	
序號	0520-362		0520-292		1504-0305		1405-0193		1215-0006		1213-0007	
備用	0850-0835		0520-293		1504-0339		1213-0010		1330-0064		1405-0194	
範圍	2~8°C		2~8°C		2~8°C		負15~負25°C		≤負70°C		15~25°C	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2024/10/1	2.9	4.4	3.9	4.4	3.7	4.2	-21.2	-17.7	-80.9	-79.4	19	19.3
2024/10/2	3.9	4.3	3.9	4.4	3.7	4.2	-21.2	-20.9	-80.9	-79.4	18.9	19
2024/10/3	3.9	4.3	3.9	4.4	3.7	4.2	-21.4	-21.1	-80.9	-79.4	18.9	19
2024/10/4	3.9	4.6	3.9	4.6	3.7	4.5	-21.2	-20.6	-80.8	-79.5	18.9	19.4
2024/10/5	4.2	4.4	4.2	4.5	3.6	4.2	-20.9	-20.8	-80.8	-79.5	18.9	19
2024/10/6	4.2	4.4	4.2	4.5	3.6	4.2	-20.9	-20.8	-80.8	-79.5	18.8	18.9
2024/10/7	4.2	5.5	4.2	4.6	3.6	4.2	-21.2	-20.7	-80.8	-79.4	18.8	19.4
2024/10/8	4.2	4.9	3.8	4.6	3.4	4.2	-21.2	-20.7	-80.9	-79.4	19	19.4
2024/10/9	3.9	5	3.8	4.5	3.3	4.1	-20.9	-20.7	-80.9	-79.4	19	21.4
2024/10/10	4.1	4.4	4	4.4	3.4	4	-20.9	-20.6	-80.9	-79.4	18.9	20.3
2024/10/11	4.1	4.8	4	4.5	3.5	4.3	-20.9	-20.4	-80.9	-78.8	18.8	19.6
2024/10/12	4.1	4.4	4	4.4	3.5	4.1	-20.7	-20.4	-80.8	-79.4	18.8	19.1
2024/10/13	4.1	4.4	4	4.4	3.6	4.2	-20.7	-20.4	-80.8	-79.2	18.7	18.9
2024/10/14	4.1	5.6	3.9	4.6	3.5	4.2	-21.1	-20.4	-80.8	-79.5	18.7	19.3
2024/10/15	4.2	4.9	3.8	4.4	3.6	4.8	-21.2	-20.5	-81.2	-76.1	18.8	19.3
2024/10/16	4	5.6	3.6	4.4	3.8	4.8	-21	-20.4	-81.1	-78.5	18.8	19.6
2024/10/17	3.9	5	3.8	4.4	3.9	4.3	-21	-20.5	-81.1	-79.4	18.8	19.3
2024/10/18	3.9	5.1	3.8	4.3	3.9	4.4	-21.1	-20.6	-81.1	-78.9	18.9	19.3
2024/10/19	3.9	4.3	3.8	4.2	3.9	4.4	-20.8	-20.6	-81	-79.4	18.9	19
2024/10/20	3.8	4.3	3.8	4.3	3.9	4.4	-20.9	-20.6	-80.9	-79.1	18.7	18.9
2024/10/21	3.9	4.8	3.8	4.3	3.9	4.6	-21.4	-20.6	-80.9	-79	18.7	19.3
2024/10/22	4.1	4.6	3.7	4.3	4	4.7	-21.4	-20.7	-80.9	-79.4	18.9	19.6
2024/10/23	3.9	4.5	3.8	4.3	4	4.7	-20.9	-20.4	-81.1	-79.4	18.9	19.3
2024/10/24	3.9	4.4	3.8	4.3	4	4.4	-21.2	-20.6	-80.9	-77.7	18.8	21.3
2024/10/25	4	4.4	3.9	4.5	3.9	4.4	-20.9	-20.5	-80.8	-78.4	19.3	20.3
2024/10/26	4.1	4.3	4.1	4.4	3.9	4.4	-20.8	-20.6	-80.8	-79.4	19.2	19.4
2024/10/27	4.1	4.3	4.1	4.4	3.9	4.4	-20.8	-20.6	-80.8	-77.9	18.9	19.1
2024/10/28	4.1	4.7	4.1	4.4	3.9	4.4	-20.8	-20.5	-80.7	-75.4	18.8	19.4
2024/10/29	4.1	4.8	3.7	4.4	3.9	4.5	-20.9	-20.6	-79.9	-78.1	18.8	19.5
2024/10/30	3.4	4.9	3.8	4.4	4.1	4.5	-20.9	-20.2	-79.9	-78.4	19	19.7
2024/10/31	3.4	4.2	3.9	4.4	4.1	4.5	-20.4	-20.2	-79.9	-77.9	19.3	20.7

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師: 藥學部 邱保祥 113
 師(三)級藥師 1101
 藥學部 邱子芹 113
 師(三)級藥師 1104

總藥師: 藥學部 廖志峰 113
 師(一)級藥師 1154
 1150

業管主管: 藥學部 何沁沁 1106
 科主任 1140

溫度均在正常範圍內