

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 113 年 10 月

	零號冰箱		四號冰箱		五號冰箱		八號冰箱		九號冰箱		室溫2	
序號	2052-0128		1611-0062		1715-0118		2123-0024		2123-0028		1715-0116	
備用	2052-0129		1611-0052		1715-0119		2123-0025		2123-0029		1715-0117	
範圍	2~8°C		2~8°C		2~8°C		2~8°C		2~8°C		15~25°C	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2024/10/1	3.8	4.4	4.1	4.8	3.9	4.6	4.7	5.2	5.1	5.6	20.2	20.9
2024/10/2	3.7	4.4	4.1	4.7	3.9	4.4	4.7	5.2	4.9	5.5	20.3	20.6
2024/10/3	3.7	4.4	4.1	4.7	3.9	4.4	4.7	5.2	5	5.5	20.4	20.6
2024/10/4	3.6	4.4	4	5.3	3.9	4.7	4.7	5.3	5.1	5.7	20.5	21.1
2024/10/5	3.6	4.3	4.2	4.8	3.9	4.4	4.7	5.2	5.2	5.6	20.5	20.7
2024/10/6	3.6	4.3	4.2	4.8	3.9	4.4	4.7	5.2	5.2	5.6	20.5	20.6
2024/10/7	3.6	4.3	3.8	4.9	3.9	4.7	4.7	5.3	5.2	5.6	20.4	21.1
2024/10/8	3.6	4.3	3.7	4.6	3.9	4.4	4.6	5.2	5	5.6	20.6	21.1
2024/10/9	3.6	4.3	3.6	4.8	3.7	4.6	4.6	5.1	4.9	5.7	20.6	21.8
2024/10/10	3.6	4.3	3.7	4.4	3.8	4.4	4.7	5.1	5.2	5.6	20.7	21.7
2024/10/11	3.6	4.5	3.7	4.6	3.9	4.8	4.7	5.2	5.1	6.1	20.5	20.9
2024/10/12	3.6	4.4	3.9	4.6	3.9	4.4	4.7	5.1	5.1	5.6	20.5	20.7
2024/10/13	3.6	4.4	3.9	4.6	3.9	4.4	4.7	5.1	5.1	5.6	20.3	20.6
2024/10/14	3.6	4.4	3.5	4.7	3.9	4.6	4.7	5.1	5.1	5.6	20.2	20.6
2024/10/15	3.7	4.4	3.5	4.6	3.9	4.5	4.7	5.3	4.9	5.7	20.3	20.7
2024/10/16	3.8	4.5	3.6	4.7	3.8	4.7	4.7	5.2	4.8	5.6	20.3	20.7
2024/10/17	3.7	4.5	3.6	4.5	3.5	4.9	4.7	7.1	4.7	5.4	20.4	20.8
2024/10/18	3.6	4.5	3.5	4.5	3.6	4.4	4.7	5.3	4.5	5.8	20.5	20.8
2024/10/19	3.6	4.4	3.7	4.5	3.6	4.3	4.7	5.1	5.3	5.7	20.4	20.6
2024/10/20	3.6	4.4	3.7	4.6	3.6	4.3	4.7	5.1	5.3	5.7	20.2	20.4
2024/10/21	3.6	4.4	3.6	4.7	3.6	4.3	4.7	5.1	5.4	5.9	20.2	20.6
2024/10/22	3.8	4.5	3.6	5	3.5	4.6	4.5	5.2	5.4	6.3	20.4	20.8
2024/10/23	3.8	4.4	3.8	4.7	3.7	4.4	4.5	5.3	5.4	5.9	20.5	20.7
2024/10/24	3.9	4.4	3.9	4.8	3.6	4.4	4.9	5.4	5.4	5.9	20.5	21.3
2024/10/25	3.8	4.4	3.9	4.7	3.6	4.4	4.7	5.3	5.3	5.9	20.8	21.3
2024/10/26	3.8	4.3	3.9	4.7	3.7	4.3	4.7	5.1	5.5	5.9	20.7	21.1
2024/10/27	3.8	4.3	3.9	4.7	3.7	4.3	4.7	5.1	5.5	5.9	20.5	20.7
2024/10/28	3.8	4.3	3.9	4.8	3.7	5.5	4.8	5.2	5.4	5.8	20.3	20.7
2024/10/29	3.8	4.4	3.7	4.7	3.7	4.7	4.7	5.1	5.3	5.8	20.5	20.9
2024/10/30	3.8	4.4	3.9	4.8	3.4	4.4	4.7	5.3	5.4	5.8	20.6	20.9
2024/10/31	3.8	4.4	4	4.8	3.3	4.2	4.7	5.1	5.4	5.9	20.8	21.3

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師 **藥學部 邱保祥** 113
藥學部 邱子芹 113
 (三)級藥師 101
 (三)級藥師 104

總藥師: **藥學部 廖志峰** 165

業管主管: **藥學部 何沁沁** 1106
 科主任 1430

溫度均在正常範圍內