

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 114年 04月

	一號冰箱		二號冰箱		三號冰箱		負20度冰箱		負80度冰箱		室溫1	
序號	0520-362		0520-292		1504-0305		1405-0193		1215-0006		1213-0007	
備用	0850-0835		0520-293		1504-0339		1213-0010		1330-0064		1405-0194	
範圍	2~8℃		2~8℃		2~8℃		負15~負25℃		≤負70℃		15~25℃	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2025/4/1	4	4.5	3.9	4.3	3.9	4.5	-18.2	-21.9	-81.6	-79.8	18.6	19.3
2025/4/2	3.5	4.6	3.9	4.3	3.9	4.6	-20.8	-21.2	-81.7	-79.8	18.6	19
2025/4/3	4.1	4.6	3.9	4.3	3.9	4.4	-19.8	-21.4	-81.7	-79.8	18.6	18.8
2025/4/4	4.2	4.6	3.9	4.3	3.9	4.4	-20.8	-21	-81.7	-79.3	18.2	18.8
2025/4/5	4.1	4.6	3.9	4.3	3.9	4.4	-20.9	-21	-81.6	-79.7	18.2	18.2
2025/4/6	4.1	4.6	3.9	4.2	3.9	4.5	-20.9	-21	-81.7	-79.1	18.1	18.2
2025/4/7	3.9	4.8	3.9	4.3	3.9	4.5	-20.7	-21.2	-81.7	-79.3	18	18.7
2025/4/8	3.9	4.8	3.8	4.6	3.9	4.5	-20.8	-21.2	-81.7	-79	18.2	18.8
2025/4/9	3.7	4.6	3.9	4.3	3.5	4.4	-19.8	-21.2	-81.7	-79.2	18.3	18.8
2025/4/10	4.2	5.3	3.8	4.3	3.5	4.4	-18.2	-22.5	-81.6	-79.5	18.4	18.9
2025/4/11	3.9	4.7	3.8	4.2	3.6	4.2	-19.6	-22.4	-81.5	-79.1	18.2	19.1
2025/4/12	4.2	4.6	3.9	4.1	3.6	4.2	-20.9	-21.4	-81.5	-79.7	18.6	18.7
2025/4/13	4.2	4.6	3.9	4.2	3.6	4.2	-20.8	-21.2	-81.5	-79.4	18	18.6
2025/4/14	4.1	5.7	3.9	4.2	3.2	4.6	-20.8	-21.6	-81.5	-79.2	18.4	19.5
2025/4/15	3.7	4.9	3.9	4.3	3.1	4.2	-20.7	-21.4	-81.6	-78.9	19.1	19.6
2025/4/16	3.6	4.8	3.9	4.3	3.1	4.4	-20.3	-21.2	-81.8	-78.6	19.2	19.7
2025/4/17	3.3	4.4	4	4.3	3.3	4.1	-17.1	-22	-81.2	-78.8	19.3	19.8
2025/4/18	3.2	4	4	4.3	3.3	4.4	-21	-22	-81.2	-79.4	19.4	20.2
2025/4/19	3.1	4	4	4.3	3.8	4.3	-20.7	-22.5	-81.2	-79.4	19.5	19.7
2025/4/20	3.3	4	4	4.3	3.9	4.3	-22	-22.4	-81.2	-78.6	19.5	19.6
2025/4/21	3.3	4.3	4	4.7	3.8	4.4	-21.7	-22.3	-81.2	-79.3	19.4	20.1
2025/4/22	3.6	4.4	3.9	4.6	3.9	4.4	-20.3	-22.7	-81.2	-78.6	19.7	20.4
2025/4/23	3.1	4.3	3.9	4.5	3.9	4.3	-21	-23.4	-81.2	-78.6	19.7	20.2
2025/4/24	3.1	6.3	3.9	4.3	3.7	4.7	-21.6	-22.4	-81.7	-78.7	19.6	20.1
2025/4/25	3	4.1	3.9	4.3	3.7	4.2	-21.7	-22.4	-81.2	-79.4	19.5	19.9
2025/4/26	3.3	4.1	3.9	4.3	3.7	4.2	-21.8	-22.2	-81.2	-79.4	19.4	19.6
2025/4/27	3.3	4.1	3.9	4.3	3.7	4.2	-21.9	-22.2	-81.2	-79.1	19.4	19.4
2025/4/28	3.2	6.2	3.9	4.4	3.8	4.5	-21.2	-22.4	-81.4	-78.4	19.4	19.8
2025/4/29	3.6	4.6	3.9	4.3	3.7	4.2	-21.6	-22.4	-81.2	-79.3	19	19.6
2025/4/30	2.9	4.3	3.9	4.5	3.6	4.2	-21.7	-22.7	-81.2	-76.9	19.2	19.9

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

溫度均在正常範圍內

承辦藥師:

藥學部 莊高文
契約醫務管理師

藥學部 邱保祥
師(三)級藥師

總藥師:

藥學部 廖志峰
師(二)級藥師

業管主管:

藥學部 何沁沁
科主任