

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 114 年 04 月

	零號冰箱		四號冰箱		五號冰箱		八號冰箱		九號冰箱		室溫2	
序號	2052-0128		1611-0062		1715-0118		2123-0024		2123-0028		1715-0116	
備用	2052-0129		1611-0052		1715-0119		2123-0025		2123-0029		1715-0117	
範圍	2~8℃		2~8℃		2~8℃		2~8℃		2~8℃		15~25℃	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2025/4/1	4.1	4.7	4.7	5.2	4.1	5.2	4.4	5.1	5.9	6.4	19.6	19.8
2025/4/2	4.4	4.8	4.7	5.2	3.9	4.8	4.4	5.2	5.8	6.6	19.5	19.7
2025/4/3	4.4	4.7	4.7	5.1	4	4.7	4.4	5.1	6.4	6.6	19.4	19.6
2025/4/4	4.4	4.7	4.7	5.1	4	4.8	4.4	5.2	6.4	6.6	19.3	19.6
2025/4/5	4.4	4.7	4.7	5.1	4	4.8	4.4	5.2	6.4	6.5	19.2	19.3
2025/4/6	4.4	4.7	4.7	5.1	4	4.8	4.4	5.2	6.4	6.4	19.1	19.2
2025/4/7	4.1	4.8	4.7	5.1	3.3	4.9	4.4	5.2	6.4	6.6	18.9	19.4
2025/4/8	4	4.5	4.5	5.1	3.5	4.9	4.4	5.2	5.6	6.7	19.4	19.8
2025/4/9	3.9	4.4	4.6	5.1	4	5.2	4.4	5.2	5.4	5.9	19.4	19.9
2025/4/10	4.2	4.8	4.7	5.1	4.2	5.1	4.4	5.1	4.9	5.8	19.4	19.9
2025/4/11	4.2	4.8	4.5	5.2	4.4	5.2	4.5	5.2	4.9	5.3	19.6	20
2025/4/12	4.6	4.7	4.5	4.9	4.3	5.1	4.9	5.2	4.9	5.2	19.8	19.9
2025/4/13	4.5	4.7	4.6	4.9	4.3	5.1	4.9	5.2	4.9	5.2	19	19.7
2025/4/14	4.5	4.9	4.6	5.3	4.3	5.3	4.9	5.4	4.9	5.3	19.2	20.1
2025/4/15	3.6	5.1	4.8	5.3	4.1	5.1	4.7	5.3	4.9	5.3	20.1	20.4
2025/4/16	3.3	4	4.8	5.3	3.9	4.9	4.7	5.6	4.8	5.2	20.3	20.7
2025/4/17	3.3	4	4.7	5.6	3.9	4.7	4.7	5.1	4.9	5.3	20.5	20.9
2025/4/18	3.7	4.3	4.7	5.2	3.6	4.7	4.7	5.1	4.7	6.2	20.7	20.9
2025/4/19	3.9	4.1	4.7	5.2	4	4.7	4.7	5.1	5.2	5.4	20.6	20.8
2025/4/20	3.9	4.1	4.7	5.2	3.9	4.7	4.7	5.1	5.2	5.4	20.6	20.6
2025/4/21	3.7	4.1	4.7	5.4	3.6	5.3	4.7	5.1	5.2	5.4	20.5	21.1
2025/4/22	3.7	4.2	4.9	5.4	4.4	5.1	4.7	5.1	4.9	5.4	20.8	21.3
2025/4/23	3.9	4.2	4.9	5.5	4.4	5.3	4.7	5.2	4.9	5.4	20.9	21.3
2025/4/24	3.8	4.2	4.4	6.1	4.3	5.1	4.7	5.1	5	5.7	20.9	21.1
2025/4/25	3.9	4.2	4.4	5.1	4.3	5.4	4.5	5.1	5.2	5.7	20.7	21.1
2025/4/26	3.9	4.2	4.6	5.1	4.3	4.9	4.5	5	5.2	5.5	20.7	20.9
2025/4/27	3.9	4.2	4.6	5.1	4.3	4.9	4.5	5	5.2	5.5	20.6	20.7
2025/4/28	3.9	4.4	4.6	5.3	4.3	5.1	4.6	5	5.2	5.7	20.6	20.9
2025/4/29	4.1	4.7	4.6	5.3	4.2	5.1	4.4	5	5.3	5.7	20.5	20.7
2025/4/30	4	4.4	4.7	5.6	4.2	4.9	4.4	4.9	5.3	5.6	20.4	20.8

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

溫度均在正常範圍內

承辦藥師:

藥學部葉馨 114 0501
藥學部邱保祥 114 0501
藥學部邱保祥 114 0501

總藥師:

藥學部廖志峰 114 0502
藥學部廖志峰 114 0502
藥學部廖志峰 114 0502

業管主管:

藥學部何沁沁 114 0502
藥學部何沁沁 114 0502
藥學部何沁沁 114 0502