

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 114 年 05 月

	零號冰箱		四號冰箱		五號冰箱		八號冰箱		九號冰箱		室溫2	
序號	2052-0128		1611-0062		1715-0118		2123-0024		2123-0028		1715-0116	
備用	2052-0129		1611-0052		1715-0119		2123-0025		2123-0029		1715-0117	
範圍	2~8℃		2~8℃		2~8℃		2~8℃		2~8℃		15~25℃	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2025/5/1	4.3	5.2	4.7	6	4.2	4.9	4.5	5.1	5.3	6.2	20.8	21.1
2025/5/2	3.6	4.5	4.7	5.6	4.2	4.7	4.5	5.1	5.3	5.7	20.6	21.1
2025/5/3	3.6	3.8	4.7	5.1	4.3	4.7	4.4	5.1	5.3	5.7	20.6	20.8
2025/5/4	3.6	3.8	4.7	5.1	4.3	4.7	4.5	5	5.3	5.7	20.4	20.6
2025/5/5	3.6	3.8	4.7	5.7	3.7	4.7	4.5	5.1	5.3	5.7	20.3	20.7
2025/5/6	3.6	3.9	4.8	5.3	4.2	4.7	4.5	5.1	5.3	5.7	20.6	21.1
2025/5/7	3.6	3.8	4.7	5.7	4.2	4.7	4.6	5.2	5.3	5.7	20.7	21.1
2025/5/8	3.6	4.7	4.6	5.4	4.2	4.7	4.6	5.2	5.1	5.7	20.7	21.1
2025/5/9	3.7	3.9	4.6	5.1	4.2	4.7	4.4	5.2	5.4	5.8	20.7	21.1
2025/5/10	3.7	3.9	4.6	4.9	4.2	4.7	4.4	5	5.4	5.7	20.7	21.1
2025/5/11	3.7	3.9	4.6	4.9	4.2	4.7	4.4	5	5.4	5.7	20.4	20.7
2025/5/12	3.7	3.9	4.6	5.1	4.2	4.7	4.3	5	5.4	5.7	20.3	20.6
2025/5/13	3.5	4	4.6	5	4.2	4.9	4.2	4.9	5.3	5.7	20.4	20.9
2025/5/14	3.5	3.7	4.6	5.2	4.2	4.8	4.1	4.9	5.2	5.6	20.6	21.1
2025/5/15	3.5	4.3	4.7	6	4.3	4.9	4	4.4	4.9	5.6	20.8	21.3
2025/5/16	4.2	4.3	4.6	5.3	4.3	4.8	4	4.4	4.7	5.6	20.8	21.2
2025/5/17	4.2	4.3	4.6	5	4.3	4.8	3.9	4.4	4.7	5.1	20.3	20.8
2025/5/18	4.2	4.2	4.6	5.1	4.3	4.8	3.9	4.3	4.7	5.1	20.3	20.6
2025/5/19	4	4.4	4.3	5.2	4.3	4.8	3.9	4.3	4.4	5.2	20.5	20.9
2025/5/20	3.9	4.2	4.2	4.8	3.9	4.7	3.9	4.3	4.4	4.9	20.7	21.1
2025/5/21	3.9	4.2	4.2	4.8	4.2	4.8	3.9	4.3	4.4	5.1	20.8	21.3
2025/5/22	3.9	4.6	4.2	5	4.2	4.8	3.9	4.3	4.5	5.2	20.9	21.4
2025/5/23	3.9	4.4	4	4.8	4.2	4.9	3.9	4.5	4.7	5.9	20.7	21.2
2025/5/24	3.9	4.1	4	4.7	4.2	4.7	4.2	4.5	5.2	5.6	20.8	21.1
2025/5/25	3.9	4.1	4.1	4.7	4.2	4.7	4.3	4.6	5.2	5.7	20.5	20.8
2025/5/26	3.9	4.3	4.1	4.7	4.1	4.7	4.3	4.5	5.1	5.7	20.4	20.7
2025/5/27	3.9	4.1	4.1	4.7	4.1	4.7	4.3	4.6	5.2	5.8	20.4	20.8
2025/5/28	3.9	4.2	4	5.2	4	4.8	4.3	4.6	5.2	6	20.6	20.9
2025/5/29	3.9	4.2	4	4.6	4.1	4.9	4.3	4.6	5.7	6.4	20.8	21.1
2025/5/30	3.9	4	4.1	4.6	4.1	4.7	4.3	4.6	5.8	6.1	20.6	20.9
2025/5/31	3.9	4	4.1	4.6	4.1	4.7	4.3	4.6	5.8	6.1	20.5	20.6

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

溫度均在正常範圍內

承辦藥師:

藥學部 葉馨
契約醫務管理組
藥學部 邱保祥
師(三)級藥師

總藥師:

藥學部 廖志峰
師(二)級藥師

業管主管:

藥學部 何沁沁
科主任