

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 114年 06月

	一號冰箱		二號冰箱		三號冰箱		負20度冰箱		負80度冰箱		室溫1	
序號	0520-362		0520-292		1504-0305		1405-0193		1215-0006		1213-0007	
備用	0850-0835		0520-293		1504-0339		1213-0010		1330-0064		1405-0194	
範圍	2~8℃		2~8℃		2~8℃		負15~負25℃		≤負70℃		15~25℃	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2025/6/1	4	4.4	3.5	3.8	3.6	4	-21.7	-21.7	-78.5	-81.5	19.2	19.2
2025/6/2	3.8	6.3	3.5	3.9	3.4	4.2	-22.2	-21.5	-81.5	-78.7	19.4	19.7
2025/6/3	3.6	4.6	3.5	4.2	3.2	4.9	-21.9	-19.2	-81.5	-79.4	19.4	19.8
2025/6/4	3.6	4.4	3.7	4.1	3.8	4.4	-22.1	-16.3	-81.5	-79.4	19.4	19.8
2025/6/5	3.2	4.5	3.8	4.1	3.8	4.3	-22.2	-20.6	-81.5	-79.4	17.9	19.8
2025/6/6	3.3	4.3	3.7	4.1	3.7	4.4	-22.9	-16.3	-81.5	-79.4	19.2	19.9
2025/6/7	3.3	4	3.7	4	3.7	4.2	-22.7	-22.6	-81.5	-79.4	19.2	19.4
2025/6/8	3.3	4	3.8	4	3.8	4.3	-22.7	-22.6	-81.5	-79.3	19.2	19.2
2025/6/9	3.3	4.7	3.7	4.1	3.5	4.8	-22.7	-20	-81.5	-79.1	19.1	19.7
2025/6/10	2.8	4.1	3.6	4	3.4	4.1	-22.5	-21.3	-81.5	-79	19.3	19.8
2025/6/11	2.8	4.9	3.7	4.1	3.3	4.1	-23.4	-21.4	-82	-79	19.4	20.2
2025/6/12	3.4	4.4	3.7	4.3	3.3	4.7	-23.2	-21.7	-81.7	-79.3	19.5	20.1
2025/6/13	3.4	4.5	3.6	4.4	3.5	4.9	-22.5	-21.8	-81.2	-76.9	19.9	21.1
2025/6/14	4	4.4	3.6	3.9	3.5	4.1	-22.2	-22.1	-79.2	-77.6	19.9	20.2
2025/6/15	4	4.4	3.7	3.8	3.5	4.1	-22.4	-22	-79.2	-77.5	19.7	19.9
2025/6/16	3.1	4.6	3.7	4	3.5	4.1	-22.9	-21.7	-79.4	-77.5	19.6	20.2
2025/6/17	2.8	4.6	3.7	4.1	3.4	4.2	-22.7	-22.1	-79.4	-77.6	19.5	20
2025/6/18	3.9	4.7	3.7	4	3.4	4.1	-23.5	-21.7	-79.3	-77.6	19.5	20.1
2025/6/19	4	4.7	3.7	4.1	3.4	4.1	-22.9	-20.9	-79.4	-77.6	19.7	20.1
2025/6/20	4	4.6	3.8	4.1	3.4	4.7	-22.7	-22.3	-79.3	-77.6	19.2	20.1
2025/6/21	4	4.5	3.7	4.1	3.6	4.1	-22.9	-22.6	-79.2	-77.6	18.8	19.2
2025/6/22	4	4.5	3.8	4	3.6	4.1	-22.9	-22.7	-79.2	-77.6	18.6	18.8
2025/6/23	4	4.5	3.8	5.5	3.6	4.2	-22.9	-22.1	-79.2	-77.6	18.6	19.3
2025/6/24	3.9	4.6	4.9	5.3	3.5	4.2	-22.9	-22.2	-79.2	-77.6	18.6	19.2
2025/6/25	3.9	5.3	5.2	5.6	3.5	4.2	-22.9	-21.4	-79.2	-77.6	18.6	19.3
2025/6/26	3.6	4.3	4.9	6	3.5	4.2	-23	-22	-79.3	-77.6	18.8	19.4
2025/6/27	3.4	4.2	4.9	6.1	3.6	4.1	-22.9	-22.3	-79.3	-77.6	18.9	19.5
2025/6/28	3.9	4.1	5.2	5.2	3.6	4	-22.9	-22.7	-79.2	-77.6	18.7	18.9
2025/6/29	3.9	4.2	5.1	5.2	3.6	4.1	-22.9	-22.6	-79.2	-77.6	18.6	18.6
2025/6/30	3.9	4.2	5.1	5.6	3.5	4.1	-22.9	-22.6	-79.2	-77.6	18.4	18.9

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

溫度均在正常範圍內

承辦藥師:

藥學部 葉馨
契約醫務管理組員

1400701

藥學部 邱保祥
師(三)級藥師

總藥師:

藥學部 陳奇良
師(二)級藥師

業管主管:

藥學部 李婉詩
師(二)級藥師