

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 114 年 06 月

	零號冰箱		四號冰箱		五號冰箱		八號冰箱		九號冰箱		室溫2	
序號	2052-0128		1611-0062		1715-0118		2123-0024		2123-0028		1715-0116	
備用	2052-0129		1611-0052		1715-0119		2123-0025		2123-0029		1715-0117	
範圍	2~8℃		2~8℃		2~8℃		2~8℃		2~8℃		15~25℃	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2025/6/1	3.9	4	4.4	4.9	4.4	4.9	4.4	4.9	5.7	6	20.3	20.4
2025/6/2	4.7	4.9	4.4	5.2	4.3	4.9	4.4	4.9	5.6	6.2	20.3	20.8
2025/6/3	4.7	4.9	4.4	5.1	4.3	5	4.4	4.9	5.6	5.9	20.7	21.1
2025/6/4	3.8	4.9	4.4	4.9	4.3	4.8	4.5	4.9	5.4	6	20.6	21.1
2025/6/5	3.8	4.3	4.4	4.9	4.2	4.8	4.5	5.1	5.4	6	20.6	20.9
2025/6/6	3.9	4.2	4.3	4.8	4.2	5	4.6	4.9	5.4	5.8	20.6	20.9
2025/6/7	3.9	4.1	4.4	4.8	4.2	4.8	4.6	4.9	5.5	5.8	20.5	20.7
2025/6/8	3.9	4.1	4.4	4.8	4.2	4.8	4.6	4.9	5.4	5.8	20.3	20.5
2025/6/9	3.9	4.8	4.4	4.9	4.2	4.9	4.7	5	4.6	5.7	20.2	20.7
2025/6/10	4.2	4.6	4.4	5.1	4.2	4.9	4.5	4.9	4.6	5.3	20.4	20.9
2025/6/11	4.2	4.4	4.5	5.1	4.4	5.1	4.5	4.9	4.7	5.4	20.8	21.3
2025/6/12	3.9	4.4	4.5	5.1	3.7	5.3	4.5	4.9	4.7	5.5	20.9	21.3
2025/6/13	3.9	4.4	4.4	5.3	4.3	5.3	4.4	4.9	4.9	5.7	21.1	21.4
2025/6/14	3.9	4.1	4.5	4.9	4.3	4.8	4.5	4.9	5.1	5.6	21.1	21.3
2025/6/15	3.9	4.2	4.5	4.9	4.3	4.8	4.5	4.9	5.1	5.6	20.7	21.1
2025/6/16	3.9	4.4	4.5	5.2	4.3	4.8	4.4	4.9	4.3	6.1	20.6	21.1
2025/6/17	4.2	4.4	4.6	5.1	4.2	4.7	4.4	4.8	4.3	4.9	20.7	21.1
2025/6/18	3.9	4.5	4.5	5.1	4.2	4.8	4.4	4.9	4.4	5.2	20.7	21.1
2025/6/19	3.9	4.3	4.5	5.1	4.2	4.8	4.4	4.9	4.4	5.4	20.9	21.2
2025/6/20	3.9	4.7	4.5	5.2	4.2	4.7	4.3	4.8	4.4	5.2	20.7	21.2
2025/6/21	4.3	4.4	4.5	4.9	4.2	4.7	4.4	4.9	4.4	5.1	20.1	20.7
2025/6/22	4.3	4.4	4.5	4.9	4.2	4.7	4.4	4.9	4.4	5.1	19.9	20.1
2025/6/23	4.2	4.7	4.4	5.1	4.2	4.7	4.4	5.5	4.3	5.1	19.7	20.2
2025/6/24	4.1	4.4	4.4	5.1	3.9	4.8	4.3	4.9	4.3	5.1	19.9	20.3
2025/6/25	3.9	4.9	4.4	5.1	4.2	5.3	4.4	4.9	4.2	5.3	20.1	20.4
2025/6/26	3.9	4.4	4.4	4.9	4	5.3	4.3	4.8	4.1	4.9	20.3	20.7
2025/6/27	3.9	4.6	4.5	4.9	3.9	5.3	4.4	4.8	4.2	5.1	20.3	20.6
2025/6/28	3.9	4.1	4.5	4.9	3.9	4.5	4.4	4.9	4.2	4.8	19.8	20.3
2025/6/29	3.9	4.1	4.5	4.9	3.9	4.6	4.4	4.9	4.2	4.8	19.7	19.8
2025/6/30	3.9	4.3	4.5	4.9	3.9	4.8	4.4	4.9	4.2	5.1	19.5	20.1

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

溫度均在正常範圍內

承辦藥師:

藥學部 葉馨
契約醫務管理組員
1140601

總藥師:

藥學部 陳奇良
師(二)級藥師
0702
400

業管主管:

藥學部 李婉詩
師(二)級藥師
0703

藥學部 邱保祥
師(三)級藥師
114
0101